



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN 1060-4:2004

Non-invasive sphygmomanometers - Part 4: Test procedures to determine the overall system accuracy of automated non-invasive

Tensiomètres non invasifs - Partie 4 :
Procédures pour déterminer la précision
de l'ensemble du système des
tensiomètres non invasifs automatiques

Nichtinvasive Blutdruckmessgeräte - Teil
4: Prüfverfahren zur Bestimmung der
Messgenauigkeit von automatischen
nichtinvasiven Blutdruckmessgeräten

National Foreword

This European Standard EN 1060-4:2004 was adopted as Luxembourgish Standard ILNAS-EN 1060-4:2004.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

EUROPEAN STANDARD ILNAS-EN 1060-4:2004
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 1060-4

September 2004

ICS 11.040.55

English version

Non-invasive sphygmomanometers - Part 4: Test procedures to determine the overall system accuracy of automated non-invasive sphygmomanometers

Tensiomètres non invasifs - Partie 4 : Procédures pour déterminer la précision de l'ensemble du système des tensiomètres non invasifs automatiques

Nichtinvasive Blutdruckmessgeräte - Teil 4: Prüfverfahren zur Bestimmung der Messgenauigkeit von automatischen nichtinvasiven Blutdruckmessgeräten

This European Standard was approved by CEN on 9 July 2004.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

| | |
|---|----|
| Foreword | 4 |
| 1 Scope | 5 |
| 2 Normative references | 5 |
| 3 Terms and definitions | 5 |
| 4 Requirements | 6 |
| 4.1 General information on the non-invasive reference methods | 6 |
| 4.2 General information on the invasive reference methods | 6 |
| 4.3 Selection of clinical test method | 6 |
| 4.4 Application of test method | 7 |
| 4.4.1 Non-invasive reference measurement | 7 |
| 4.4.2 Invasive reference measurement | 8 |
| 4.5 Test equipment | 8 |
| 4.5.1 Non-invasive reference measurement | 8 |
| 4.5.2 Invasive blood pressure measurement | 8 |
| 4.6 Subjects | 9 |
| 4.6.1 General | 9 |
| 4.6.2 Non-invasive reference measurement | 9 |
| 4.6.3 Invasive reference measurement | 10 |
| 4.7 Cuff size | 10 |
| 4.8 Procedure | 10 |
| 4.9 Evaluation of results | 10 |
| 4.9.1 General | 10 |
| 4.9.2 Non-invasive reference measurement | 11 |
| 4.9.3 Documentation | 13 |
| 5 Non-invasive test methods | 13 |
| 5.1 Test method N1 - simultaneous blood pressure measurement on the same upper arm | 13 |
| 5.1.1 Procedure | 13 |
| 5.1.2 Evaluation | 14 |
| 5.2 Test method N2 - blood pressure measurement on the opposite arms | 14 |
| 5.2.1 Procedure | 14 |
| 5.2.2 Evaluation | 16 |
| 5.2.3 Documentation | 16 |
| 5.3 Test method N3 - sequential blood pressure measurement on the same arm | 17 |
| 5.3.1 Procedure | 17 |
| 5.3.2 Evaluation | 18 |
| 5.4 Test method N4 - simultaneous blood pressure measurement on the same upper arm at rest and under physical load | 18 |
| 5.4.1 Procedure | 18 |
| 5.4.2 Evaluation | 18 |
| 5.4.3 Documentation | 18 |
| 5.5 Test method N5 - ambulatory blood pressure measurement on the same upper arm | 18 |
| 5.5.1 Procedure of measurement | 18 |
| 5.5.2 Evaluation | 19 |
| 5.5.3 Documentation | 19 |
| 5.6 Test method N6 - ambulatory blood pressure measurement on the opposite arms | 19 |
| 5.6.1 Procedure | 19 |
| 5.6.2 Evaluation | 20 |
| 5.6.3 Documentation | 20 |
| 6 Invasive test methods | 20 |
| 6.1 Test method I 1 - invasive blood pressure measurement on adults | 20 |
| 6.1.1 Procedure | 20 |
| 6.1.2 Evaluation | 20 |

| | |
|---|-----------|
| 6.1.3 Documentation..... | 21 |
| 6.2 Test method I 2 - invasive blood pressure measurement on neonates and infants..... | 21 |
| 6.2.1 Procedure | 21 |
| 6.2.2 Evaluation..... | 21 |
| 6.2.3 Documentation..... | 22 |
| Annex A (informative) Chronological order of the test method N 3 Sequential blood pressure measurement on the same arm..... | 23 |
| Annex B (informative) Example of a record of a clinical investigation according to method N1 or N3..... | 24 |
| Annex C (informative) Summary of the WHO recommendations for auscultatory blood pressure measurement | 27 |
| Annex D (informative) Observer training and assessment etc..... | 28 |
| Annex E (informative) A-deviations | 30 |
| Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EC Medical Devices | 31 |
| Bibliography | 32 |

Foreword

This document (EN 1060-4:2004) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2005, and conflicting national standards shall be withdrawn at the latest by March 2005.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

EN 1060 consists of the following Parts under the general title “Non-invasive sphygmomanometers”:

Part 1: General requirements

Part 2: Supplementary requirements for mechanical sphygmomanometers

Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems

Part 4: Test procedures to determine the overall system accuracy of automated non-invasive sphygmomanometers

Annexes A, B, C, D, E and ZA are informative.

This document includes a Bibliography.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.