

# ILNAS

Institut luxembourgeois de la normalisation  
de l'accréditation, de la sécurité et qualité  
des produits et services

**ILNAS-EN 13795-2:2019**

## **Surgical clothing and drapes - Requirements and test methods - Part 2: Clean air suits**

Operationskleidung und -abdecktücher -  
Anforderungen und Prüfverfahren - Teil 2:  
Rein-Luft-Kleidung

Vêtements et champs chirurgicaux -  
Exigences et méthodes d'essai - Partie 2 :  
Tenues de bloc

**04/2019**



## National Foreword

This European Standard EN 13795-2:2019 was adopted as Luxembourgish Standard ILNAS-EN 13795-2:2019.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

### **THIS PUBLICATION IS COPYRIGHT PROTECTED**

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

EUROPEAN STANDARD <sup>ILNAS-EN 13795-2:2019</sup> **EN 13795-2**

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2019

ICS 11.140

Supersedes EN 13795:2011+A1:2013

English Version

## Surgical clothing and drapes - Requirements and test methods - Part 2: Clean air suits

Vêtements et champs chirurgicaux - Exigences et méthodes d'essai - Partie 2 : Tenues de bloc

Operationsbekleidung und -abdecktücher - Anforderungen und Prüfverfahren - Teil 2: Rein-Luft-Kleidung

This European Standard was approved by CEN on 24 October 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**Contents**

Page

European foreword.....	4
Introduction .....	6
1 Scope.....	7
2 Normative references.....	7
3 Terms and definitions .....	7
4 Performance requirements.....	10
5 Manufacturing and processing requirements and documentation .....	11
6 Information to be supplied with the product .....	11
6.1 Information to be supplied to the user .....	11
6.2 Information to be supplied to the processor .....	11
Annex A (normative) Testing.....	12
A.1 General.....	12
A.2 Test methods and conformance.....	12
A.2.1 Test method for evaluation of cleanliness microbial/bioburden .....	12
A.2.2 Test method for evaluation of particle release .....	12
A.2.3 Test method for evaluation of bursting strength in dry state.....	13
A.2.4 Test method for evaluation of tensile strength in dry state .....	13
A.2.5 Test method for evaluation of dry microbial penetration .....	13
A.2.6 Test method for evaluation of biocompatibility .....	13
A.3 Treatment of results .....	13
Annex B (informative) Rationales.....	15
B.1 General.....	15
B.2 Cleanliness – microbial .....	15
B.3 Particle release .....	15
B.4 Bursting strength – dry .....	16
B.5 Tensile strength – dry.....	16
B.6 Resistance to microbial penetration – dry.....	16
B.7 Labelling.....	17
B.8 Treatment of results .....	17
B.9 Flammability.....	18
B.10 Electrostatic discharge.....	18
Annex C (informative) Environmental aspects.....	19
Annex D (informative) Guidance to users for selecting products.....	20

<b>D.1</b>	<b>General .....</b>	<b>20</b>
<b>D.2</b>	<b>Performance levels .....</b>	<b>20</b>
<b>D.3</b>	<b>Functional design aspects.....</b>	<b>20</b>
<b>D.3.1</b>	<b>Size .....</b>	<b>20</b>
<b>D.3.2</b>	<b>Accessories.....</b>	<b>20</b>
<b>D.4</b>	<b>Comfort.....</b>	<b>21</b>
<b>D.4.1</b>	<b>General .....</b>	<b>21</b>
<b>D.4.2</b>	<b>Clean air suits.....</b>	<b>21</b>
<b>D.4.3</b>	<b>Practical trials.....</b>	<b>21</b>
<b>Annex E (informative) Functional design .....</b>		<b>22</b>
<b>E.1</b>	<b>General .....</b>	<b>22</b>
<b>E.2</b>	<b>Test method for measuring source strength .....</b>	<b>22</b>
<b>E.2.1</b>	<b>Dispersal chamber .....</b>	<b>22</b>
<b>E.2.2</b>	<b>Operating room .....</b>	<b>23</b>
<b>E.2.3</b>	<b>Measuring bacteria carrying airborne particles .....</b>	<b>23</b>
<b>E.2.4</b>	<b>Source strength .....</b>	<b>23</b>
<b>E.3</b>	<b>Use of source strength measurements.....</b>	<b>24</b>
<b>Annex ZA (informative) Relationship between this European standard and the essential requirements of Directive 93/42/EEC [1993 OJ L 169] aimed to be covered .....</b>		<b>26</b>
<b>Bibliography .....</b>		<b>27</b>

## European foreword

This document (EN 13795-2:2019) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2019, and conflicting national standards shall be withdrawn at the latest by October 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Together with EN 13795-1:2019, this document supersedes EN 13795:2011+A1:2013.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

EN 13795 consists of the following parts, under the general title *Surgical clothing and drapes — Requirements and test methods*:

- Part 1: *Surgical drapes and gowns*
- Part 2: *Clean air suits*

The following changes have been introduced:

- a) Restriction to the product ‘clean-air suit’ in this Part of the EN 13795 standard series (for surgical drapes and gowns see EN 13795-1);
- b) Alignment of the Standard title and the Scope;
- c) Revision of the Normative references and the Bibliography;
- d) Alignment of the Clause ‘Terms and definitions’;
- e) Revision of the performance requirements in Table 1;
- f) Movement of former Clause 5 ‘Testing’ to A.1 and editorial alignment;
- g) Revision of Clause ‘Manufacturing and processing requirements’ by adding of documentary requirements and a section for the introduction of a QM system;
- h) Enhancement and improved structuring of Clause ‘Information to be supplied by the manufacturer or processor’;
- i) Deletion of the former Annex A ‘Details of significant changes between this document and the previous edition’;
- j) Complete revision and extension of Annex A ‘Testing’ (formerly Annex B ‘Test methods’);

- k) Inclusion of a new Annex B 'Rationales' which provides precise reasons for the essential requirements of this document and which is intended for users aware of the subject of this document, but who did not join whose development;
- l) Deletion of the former Annex C 'Prevention of infection in the operating room';
- m) Inclusion of a new Annex C 'Environmental aspects';
- n) Inclusion of a new Annex D 'Guidance to users for selecting products';
- o) Inclusion of a new Annex E 'Functional design';
- p) Revision of Annex ZA on the relationship to the Medical Device Directive (93/42/EEC);
- q) Complete editorial revision.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Introduction

Clean air suits are used to minimize the spread of infective agents to patients' surgical sites and equipment, through prevention of dispersal of bacteria-carrying skin scales from the operating room staff, thereby helping to prevent post-operative surgical site infections.

The performance required of working clothes for clinical staff varies with, for example, the type and duration of the procedure, and the susceptibility of the patient to infection. In infection-prone invasive operations, a clean air suit can contribute to reduction of infection risks, in conjunction with ventilation and correct working methods.

This document is intended to assist the communication between manufacturers and third parties with regard to material or product characteristics and performance requirements.

Therefore, Annex B provides comprehensive information on characteristics, measurement of performance and performance requirements. Annex C clarifies that this document does not include environmental provisions. Annex D explains the concept of performance levels and provides guidance to users for selecting products. Annex E gives information on the impact of the design of clean air suits and the source strength concept as an evaluation means for the impact of the entire clothing (including clean air suits) on particle release.

This document focuses on Essential Requirements arising from the Medical Device Directive 93/42/EEC, which are applicable to clean air suits. The requirements and guidance in this document are expected to be of help to manufacturers and users when designing, processing, assessing and selecting products. It is the intention of this document to ensure the same level of safety from single-use and reusable clean air suits throughout their useful life.