



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN ISO 20749:2018

Dentistry - Pre-capsulated dental amalgam (ISO 20749:2017)

Zahnheilkunde - Gekapseltes
zahnärztliches Amalgam (ISO
20749:2017)

Médecine bucco-dentaire - Amalgame
dentaire en capsules prédosées (ISO
20749:2017)

09/2018



National Foreword

This European Standard EN ISO 20749:2018 was adopted as Luxembourgish Standard ILNAS-EN ISO 20749:2018.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

ILNAS-EN ISO 20749:2018

EUROPEAN STANDARD **EN ISO 20749**

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2018

ICS 11.060.10

English Version

**Dentistry - Pre-capsulated dental amalgam (ISO
20749:2017)**

Médecine bucco-dentaire - Amalgame dentaire en
capsules prédosées (ISO 20749:2017)

Zahnheilkunde - Gekapseltes zahnärztliches Amalgam
(ISO 20749:2017)

This European Standard was approved by CEN on 22 July 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

ILNAS-EN ISO 20749:2018 - Preview only Copy via ILNAS e-Shop

European foreword

The text of ISO 20749:2017 has been prepared by Technical Committee ISO/TC 106 "Dentistry" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 20749:2018 by Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2019, and conflicting national standards shall be withdrawn at the latest by March 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20749:2017 has been approved by CEN as EN ISO 20749:2018 without any modification.

Dentistry — Pre-capsulated dental amalgam

*Médecine bucco-dentaire — Amalgame dentaire en capsules
prédosées*

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	3
4.1 Package and capsule contamination	3
4.2 Chemical composition and purity of the dental amalgam alloy	3
4.3 Large particles in the dental amalgam alloy powder	3
4.4 Loss of mass from the capsule during mixing	3
4.5 The yield of amalgam from the capsule	3
4.6 Consistency of the dental amalgam from capsule to capsule	4
4.7 Properties of the dental amalgam	4
4.7.1 Creep	4
4.7.2 Dimensional changes during hardening	4
4.7.3 Compressive strength at 1 h	4
4.7.4 Compressive strength at 24 h	4
4.8 Appearance of the mixed dental amalgam before setting	4
4.9 Length tolerance for the capsule	4
5 Sampling	5
6 Test methods	5
6.1 Package and capsule contamination	5
6.1.1 Principle	5
6.1.2 Test sample	5
6.1.3 Apparatus	5
6.1.4 Procedure	5
6.1.5 Expression of the results	5
6.1.6 Report	5
6.2 Chemical composition and purity of the dental amalgam alloy	5
6.2.1 Principle	5
6.2.2 Test sample	6
6.2.3 Apparatus	6
6.2.4 Procedure	6
6.2.5 Expression of results	6
6.2.6 Report	6
6.3 Large particles in the dental amalgam alloy powder	6
6.3.1 Principle	6
6.3.2 Test sample	7
6.3.3 Apparatus	7
6.3.4 Test procedure	7
6.3.5 Expression of the results	7
6.3.6 Report	7
6.4 Loss of mass from the capsule during mixing	8
6.4.1 Principle	8
6.4.2 Test sample	8
6.4.3 Apparatus	8
6.4.4 Test procedure	8
6.4.5 Expression of the results	9
6.4.6 Report	9
6.5 Yield of amalgam from the capsule	9
6.5.1 Principle	9
6.5.2 Test sample	9