



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN 16091:2022

Liquid petroleum products - Middle distillates and fatty acid methyl ester (FAME) fuels and blends - Determination of oxidation stability by

Produits pétroliers liquides - Distillats
moyens, esters méthyliques d'acides gras
(EMAG) et leurs mélanges -
Détermination de la stabilité à

Flüssige Mineralölerzeugnisse -
Mitteldestillat- und Fettsäure-
Methylester (FAME)-Kraftstoffe und -
Mischungen - Bestimmung der

10/2022



National Foreword

This European Standard EN 16091:2022 was adopted as Luxembourgish Standard ILNAS-EN 16091:2022.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

EUROPEAN STANDARD ^{ILNAS-EN 16091:2022} **EN 16091**
NORME EUROPÉENNE
EUROPÄISCHE NORM

October 2022

ICS 75.160.20

Supersedes EN 16091:2011

English Version

Liquid petroleum products - Middle distillates and fatty acid methyl ester (FAME) fuels and blends - Determination of oxidation stability by rapid small scale oxidation test (RSSOT)

Produits pétroliers liquides - Distillats moyens, esters méthyliques d'acides gras (EMAG) et leurs mélanges - Détermination de la stabilité à l'oxydation par méthode d'oxydation accélérée à petite échelle (RSSOT)

Flüssige Mineralölerzeugnisse - Mitteldestillat- und Fettsäure-Methylester (FAME)-Kraftstoffe und -Mischungen - Bestimmung der Oxidationsstabilität mit beschleunigtem Oxidationsverfahren und kleiner Probenmenge (RSSOT)

This European Standard was approved by CEN on 19 September 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	4
1 Scope.....	5
2 Normative references.....	5
3 Terms and definitions	5
4 Principle	5
5 Reagents and materials	6
6 Apparatus	6
7 Sampling and sample handling.....	6
8 Performance check of the apparatus.....	6
9 Preparation of the apparatus	7
10 Procedure	7
11 Expression of results	8
12 Precision	8
12.1 General.....	8
12.2 Repeatability, r.....	8
12.3 Reproducibility, R.....	8
13 Test report.....	9
Annex A (normative) Apparatus description.....	10
A.1 General requirements.....	10
A.2 Assembly of apparatus.....	10
A.3 Reaction vessel	10
A.4 Screw cap	10
A.5 Electric heater	11
A.6 O-ring seals.....	11
A.7 Valves	11
A.8 Pressure sensor	11
A.9 Temperature sensor	11
A.10 Connecting pipes	11
A.11 Cooling fan.....	11
Annex B (informative) Determination of induction period.....	12
Annex C (informative) Test method and precision data at 120 °C	13
C.1 Background.....	13
C.2 Short description of the implementation of the RSSOT method using a reduced measurement temperature of 120 °C	13

C.3	Precision	13
C.3.1	General	13
C.3.2	Repeatability, r	14
C.3.3	Reproducibility, R	14
C.4	Outcome of the ILS	14
	Bibliography	15

European foreword

This document (EN 16091:2022) has been prepared by Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2023, and conflicting national standards shall be withdrawn at the latest by month April 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 16091:2011.

In comparison with the previous edition EN 16091:2011, the following technical modifications have been made:

- a) addition of informative Annex C to include work that has been executed on a modified rapid small scale oxidation method at 120 °C;
- b) revision of Clause 8: the performance check of the apparatus and the recommended calibration process have been merged into a single Clause 8.1;
- c) Annex on the calibration process has been removed;
- d) the document was revised editorially.

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

1 Scope

This document specifies a method for the determination of the oxidation stability of middle distillate fuels, fatty acid methyl ester (FAME) fuel and blends thereof, under accelerated conditions, by measuring the induction period to the specified breakpoint in a reaction vessel charged with the sample and oxygen at 140 °C.

NOTE 1 For the purposes of this document, the term “% (V/V)” is used to represent the volume fraction (φ).

NOTE 2 The induction period is used as an indication for the resistance of middle distillates, fatty acid methyl ester (FAME) fuels and blends thereof against oxidation. This correlation can vary markedly under different conditions with different FAMEs and diesel fuel blends.

NOTE 3 The presence of ignition improvers can lead to lower oxidation stability results determined by this method. It has for instance been observed that the addition of 2-ethyl hexyl nitrate (2-EHN) can reduce the measured oxidation stability values. See [6] for details.

NOTE 4 For further information on the precision data at a test temperature of 120 °C see Annex C.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3170, *Petroleum liquids - Manual sampling (ISO 3170)*

EN ISO 3171, *Petroleum liquids - Automatic pipeline sampling (ISO 3171)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

breakpoint

point in the pressure-time curve, where the pressure in the vessel has dropped 10 % below the maximum pressure achieved in the current test run

3.2

induction period

time elapsed between starting the heating procedure of the reaction vessel, charged with sample and oxygen, and the breakpoint, measured in minutes

4 Principle

At ambient temperature, a known volume of a sample is placed in a reaction vessel charged with oxygen to a pressure of 700 kPa \pm 5 kPa. The reaction vessel is heated to 140 °C. The pressure in the vessel drops as the oxygen is consumed during the oxidation of the sample. The pressure in the vessel is recorded at intervals of 1 s until the breakpoint is reached. The elapsed time from start to the breakpoint is the induction period at the test temperature of 140 °C \pm 0,5 °C.

NOTE Work has been executed on the determination of oxidation stability by a modified rapid small-scale oxidation method at 120 °C. The results thereof are given in Annex C.

5 Reagents and materials

5.1 Cleaning solvent, for the removal of oxidation residues from the test vessel, of suitable purity to leave no residue on the apparatus.

NOTE Commercially available ethanol of approximately 95 % (V/V) purity was found to be suitable.

5.2 Oxygen, extra dry (<5 mg/kg water), commercially available, with a purity of not less than 99,6 %.

5.3 Cleaning tissues, lint-free for cleaning sensitive galvanic coated surfaces without scratching.

5.4 Verification fluid, hydrocarbon fuel, any hydrocarbon fuel with sufficient stability and a known induction period may be used.

NOTE In general, verification fluid with a certified induction period is available from the manufacturer of the apparatus.

6 Apparatus

6.1 Automatically controlled oxidation tester, comprising an oxidation reaction vessel containing:

- test sample cup capable of being rapidly heated;
- pressure sensor capable of measuring pressures of 1 kPa up to 2 000 kPa with a resolution of 1 kPa;
- temperature sensor with a resolution of 0,1 °C;
- pressure and temperature recording at time intervals of 1 s.

The oxidation reaction vessel shall be fitted with filling and relief valves, meant to release the pressure, and a fan to cool the vessel from test to ambient temperature.

The requirements for the apparatus are described in detail in Annex A.

6.2 Pipette, capable of delivering 5,0 ml ± 0,1 ml.

6.3 O-ring seals, see A.6.

7 Sampling and sample handling

Unless otherwise specified, samples shall be taken as described in EN ISO 3170 or EN ISO 3171.

Collect and store samples in an opaque container to minimize exposure to light.

8 Performance check of the apparatus

8.1 Verify the correct operation of the equipment by following the procedure in Clause 10 using the verification fluid (5.4). If the verification measurement result is within the certified range then the pressure and temperature sensor are working correctly, and the reaction vessel is reaching 140 °C ± 0,5 °C within the required time of 270 s ± 30 s.

8.2 If the induction period of the verification fluid does not meet the specified time, repeat 8.1. If it fails a second time, refer to the manufacturer's instructions.