



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services


ILNAS-EN ISO 16672:2021

Ophthalmic implants - Ocular endotamponades (ISO 16672:2020)

Implants ophtalmiques - Produits de
tamponnement endoculaires (ISO
16672:2020)

Ophthalmische Implantate - Okulare
Endotamponaden (ISO 16672:2020)

10/2021



National Foreword

This European Standard EN ISO 16672:2021 was adopted as Luxembourgish Standard ILNAS-EN ISO 16672:2021.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

ILNAS-EN ISO 16672:2021

EUROPEAN STANDARD **EN ISO 16672**

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2021

ICS 11.040.70

Supersedes EN ISO 16672:2015

English Version

Ophthalmic implants - Ocular endotamponades (ISO 16672:2020)

Implants ophtalmiques - Produits de tamponnement
endoculaires (ISO 16672:2020)

Ophthalmische Implantate - Okulare Endotamponaden
(ISO 16672:2020)

This European Standard was approved by CEN on 19 November 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

ILNAS-EN ISO 16672:2021 - Preview only Copy via ILNAS e-Shop

European foreword

This document (EN ISO 16672:2021) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2022, and conflicting national standards shall be withdrawn at the latest by April 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 16672:2015.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 16672:2020 has been approved by CEN as EN ISO 16672:2021 without any modification.

Ophthalmic implants — Ocular endotamponades

Implants ophtalmiques — Produits de tamponnement endoculaires

**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2020

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Intended performance	3
5 Design attributes	3
5.1 General	3
5.2 Chemical description and contaminants	3
5.3 Density	4
5.4 Gaseous expansion	4
5.5 Interfacial tension	4
5.6 Kinematic viscosity	4
5.7 Dynamic viscosity	4
5.8 Molecular mass distribution	4
5.9 Particulates	5
5.10 Refractive index	5
5.11 Spectral transmittance	5
5.12 Surface tension	5
5.13 Vapour pressure	5
6 Design evaluation	5
6.1 General	5
6.2 Evaluation of biological safety	6
6.2.1 General	6
6.2.2 Bacterial endotoxins test	6
6.2.3 Intraocular implantation test	6
6.2.4 Ethylene oxide	6
6.3 Clinical investigation	7
7 Sterilization	7
8 Product stability	7
9 Integrity and performance of the delivery system	7
10 Packaging	8
10.1 Protection from damage during storage and transport	8
10.2 Maintenance of sterility in transit	8
11 Information supplied by the manufacturer	8
Annex A (normative) Intraocular implantation test	10
Annex B (informative) Clinical investigation	11
Annex C (informative) Method for quantifying incompletely fluorinated contaminants in perfluorocarbon liquids	14
Bibliography	16