



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN ISO 11608-5:2023

Needle-based injection systems for medical use - Requirements and test methods - Part 5: Automated functions (ISO 11608-5:2022)

Systèmes d'injection à aiguille pour
usage médical - Exigences et méthodes
d'essai - Partie 5: Fonctions automatisées
(ISO 11608-5:2022)

Kanülenbasierte Injektionssysteme zur
medizinischen Verwendung -
Anforderungen und Prüfverfahren - Teil 5:
Automatisierte Funktionen (ISO

03/2023



National Foreword

This European Standard EN ISO 11608-5:2023 was adopted as Luxembourgish Standard ILNAS-EN ISO 11608-5:2023.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

English Version

Needle-based injection systems for medical use - Requirements and test methods - Part 5: Automated functions (ISO 11608-5:2022)

Systèmes d'injection à aiguille pour usage médical -
Exigences et méthodes d'essai - Partie 5: Fonctions
automatisées (ISO 11608-5:2022)

Kanülenbasierte Injektionssysteme zur medizinischen
Verwendung - Anforderungen und Prüfverfahren - Teil
5: Automatisierte Funktionen (ISO 11608-5:2022)

This European Standard was approved by CEN on 10 March 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

ILNAS-EN ISO 11608-5:2023 - Preview only Copy via ILNAS e-Shop

European foreword

The text of ISO 11608-5:2022 has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 11608-5:2023 by Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2023, and conflicting national standards shall be withdrawn at the latest by September 2023.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11608-5:2012.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 11608-5:2022 has been approved by CEN as EN ISO 11608-5:2023 without any modification.

Needle-based injection systems for medical use — Requirements and test methods —

Part 5: Automated functions

*Systèmes d'injection à aiguille pour usage médical — Exigences et
méthodes d'essai —*

Partie 5: Fonctions automatisées



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Requirements	4
4.1 General requirements	4
4.2 Medicinal product preparation.....	5
4.3 Needle preparation.....	6
4.4 Needle hiding.....	6
4.5 Priming.....	6
4.6 Dose setting.....	6
4.7 Needle insertion.....	6
4.8 Injection depth control.....	6
4.9 Dose delivery.....	7
4.10 Recording of device functions	7
4.11 Needle retraction	7
4.11.1 Completion of dose delivery	7
4.11.2 Needle retraction distance.....	7
4.11.3 Communication of completion	7
4.12 Disabling the NIS-AUTO.....	7
4.13 Needle shielding	8
4.13.1 General	8
4.13.2 Needle shielding before injection.....	8
4.13.3 Needle shielding after injection.....	8
4.14 Needle removal from the NIS-AUTO	8
5 Test methods	8
5.1 General.....	8
5.2 Test conditions.....	9
5.3 Actuation	9
5.4 Medicinal product preparation.....	9
5.5 Needle inspection	9
5.6 Needle hiding.....	9
5.7 Priming.....	10
5.8 Needle extension	10
5.9 Injection time.....	10
5.10 Dose accuracy.....	11
5.11 Retracted position.....	11
5.12 Disabling the NIS-AUTO.....	11
5.13 Needle shielding.....	11
5.13.1 Needle shielding before and after injection	11
5.13.2 Needle shielding after free fall.....	11
6 Information supplied with the NIS-AUTO	11
Annex A (informative) Rationale for requirements	12
Annex B (informative) Example of a test method for dose accuracy at intended injection depth	14
Annex C (informative) Needle extension and intended injection depth	16
Bibliography	22