



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN ISO 13164-4:2023

Water quality - Radon-222 - Part 4: Test method using two-phase liquid scintillation counting (ISO 13164-4:2023)

Qualité de l'eau - Radon 222 - Partie 4:
Méthode d'essai par comptage des
scintillations en milieu liquide à deux
phases (ISO 13164-4:2023)

Wasserbeschaffenheit - Radon-222 - Teil
4: Verfahren mittels zweistufiger
Flüssigszintillationszählung (ISO/FDIS
13164-4:2023)

07/2023



National Foreword

This European Standard EN ISO 13164-4:2023 was adopted as Luxembourgish Standard ILNAS-EN ISO 13164-4:2023.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

English Version

Water quality - Radon-222 - Part 4: Test method using two-phase liquid scintillation counting (ISO 13164-4:2023)

Qualité de l'eau - Radon 222 - Partie 4: Méthode d'essai
par comptage des scintillations en milieu liquide à
deux phases (ISO 13164-4:2023)

This European Standard was approved by CEN on 12 June 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

ILNAS-EN ISO 13164-4:2023 - Preview only Copy via ILNAS e-Shop

European foreword

This document (EN ISO 13164-4:2023) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2024, and conflicting national standards shall be withdrawn at the latest by January 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 13164-4:2020.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 13164-4:2023 has been approved by CEN as EN ISO 13164-4:2023 without any modification.

Water quality — Radon-222 —

Part 4:

Test method using two-phase liquid scintillation counting

Qualité de l'eau — Radon 222 —

*Partie 4: Méthode d'essai par comptage des scintillations en milieu
liquide à deux phases*



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms, definitions and symbols	1
3.1 Terms and definitions	1
3.2 Symbols	2
4 Principle	3
5 Sampling	3
5.1 General	3
5.2 Sampling with source preparation “on site”	3
5.3 Sampling without “on site” source preparation	3
6 Reagents and apparatus	3
6.1 Reagents	3
6.2 Apparatus	4
7 Procedure	4
7.1 Preparation of calibration sources	4
7.2 Optimization of counting conditions	4
7.3 Detection efficiency	5
7.4 Blank sample preparation and measurement	5
7.5 Sample preparation and measurement	5
8 Quality assurance and quality control program	6
8.1 General	6
8.2 Variables that could influence the measurement	6
8.3 Instrument verification	6
8.4 Contamination	6
8.5 Interference control	7
8.6 Method verification	7
8.7 Demonstration of analyst capability	7
9 Expression of results	7
9.1 General	7
9.2 Count rate	7
9.3 Calculation of activity concentration per unit of mass	8
9.4 Combined uncertainty	8
9.5 Decision threshold	8
9.6 Detection limit	9
9.7 Probabilistically symmetric coverage interval	9
9.7.1 Limits of the probabilistically symmetric coverage interval	9
9.7.2 The shortest coverage interval	10
9.8 Calculations using the activity concentration	10
10 Test report	10
Annex A (informative) Set-up parameters and validation data	12
Bibliography	16