

August 2023

ICS 11.140

Will supersede EN 13795-2:2019

English Version

Surgical clothing and drapes - Requirements and test methods - Part 2: Clean air suits

Vêtements et champs chirurgicaux - Exigences et méthodes d'essai - Partie 2 : Tenues de bloc

Operationskleidung und -abdecktücher - Anforderungen und Prüfverfahren - Teil 2: Rein-Luft-Kleidung

This draft European Standard is submitted to CEN members for enquiry. It has been drawn up by the Technical Committee CEN/TC 205.

If this draft becomes a European Standard, CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

This draft European Standard was established by CEN in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.

Recipients of this draft are invited to submit, with their comments, notification of any relevant patent rights of which they are aware and to provide supporting documentation.

Warning : This document is not a European Standard. It is distributed for review and comments. It is subject to change without notice and shall not be referred to as a European Standard.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents		Page
European foreword		3
Introduction		4
1	Scope	5
2	Normative references	5
3	Terms and definitions	5
4	Performance requirements	8
5	Manufacturing and processing requirements and documentation	9
6	Information to be supplied with the product	9
6.1	Information to be supplied to the user	9
6.2	Information to be supplied to the processor	9
Annex A (normative) Testing		10
Annex B (informative) Rationales		13
Annex C (informative) Environmental impact		17
Annex D (informative) Guidance to users for selecting products		19
Annex E (informative) Functional design		21
Annex ZA (informative) Relationship between this European standard and General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered		25
Bibliography		27

European foreword

This document (prEN 13795-2:2023) has been prepared by Technical Committee CEN/TC 205 “Non-active medical devices”, the secretariat of which is held by DIN.

This document is currently submitted to the CEN Enquiry.

This document will supersede EN 13795-2:2019.

prEN 13795-2:2023 includes the following significant technical changes with respect to EN 13795-2:2019:

- a) Clarification of testing specifications and reporting of results;
- b) Expansion of Annex C (formerly read “Environmental aspects”) to include considerations regarding environmental impact and circular economy (now Annex C “Environmental impact”);
- c) Alignment to Regulation (EU) 2017/745 (including updated Annex ZA);
- d) Update of normative references and bibliography.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

EN 13795 consists of the following parts, under the general title *Surgical clothing and drapes — Requirements and test methods*:

- *Part 1: Surgical drapes and gowns*
- *Part 2: Clean air suits*

Introduction

Clean air suits are used to minimize the spread of infective agents to patients' surgical sites and equipment, through prevention of dispersal of bacteria-carrying skin scales from the operating room staff, thereby helping to prevent post-operative surgical site infections.

The performance required of working clothes for clinical staff varies with, for example, the type and duration of the procedure, and the susceptibility of the patient to infection. In infection-prone invasive operations, a clean air suit can contribute to reduction of infection risks, in conjunction with ventilation and correct working methods.

This document is intended to assist the communication between manufacturers and third parties with regard to material or product characteristics and performance requirements.

Therefore, Annex B provides comprehensive information on characteristics, measurement of performance and performance requirements. Annex C clarifies that this document does not include environmental provisions. Annex D explains the concept of performance levels and provides guidance to users for selecting products. Annex E gives information on the impact of the design of clean air suits and the source strength concept as an evaluation means for the impact of the entire clothing (including clean air suits) on particle release.

This document focuses on General Safety and Performance Requirements (GSPR) arising from the Medical Device Regulation (EU) 2017/745, which are applicable to clean air suits. The requirements and guidance in this document are expected to be of help to manufacturers and users when designing, processing, assessing and selecting products. It is the intention of this document to ensure the same level of safety from single-use and reusable clean air suits throughout their useful life.

1 Scope

This document specifies information to be supplied to users and third-party verifiers in addition to the usual labelling of medical devices (see EN ISO 20417 and EN ISO 15223-1), concerning manufacturing and processing requirements.

This document gives information on the characteristics of single-use and reusable clean air suits used as medical devices for clinical staff, intended to prevent the transmission of infective agents between clinical staff and patients during surgical and other invasive procedures.

This document specifies test methods for evaluating the identified characteristics of clean air suits and sets performance requirements for these products.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 139:2005,¹ *Textiles — Standard atmospheres for conditioning and testing (ISO 139:2005 + Amd. 1:2011)*

EN ISO 9073-3:2023, *Nonwovens - Test methods - Part 3: Determination of tensile strength and elongation at break using the strip method (ISO 9073-3:2023)*

EN ISO 9073-10:2004, *Textiles - Test methods for nonwovens - Part 10: Lint and other particles generation in the dry state (ISO 9073-10:2003)*

EN ISO 10993-1:2020, *Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2018, including corrected version 2018-10)*

EN ISO 11737-1:2018,² *Sterilization of medical devices — Microbiological methods — Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018)*

EN ISO 13938-1:2019, *Textiles - Bursting properties of fabrics - Part 1: Hydraulic method for determination of bursting strength and bursting distension (ISO 13938-1:2019)*

EN ISO 22612:2005, *Clothing for protection against infectious agents - Test method for resistance to dry microbial penetration (ISO 22612:2005)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp/>
- IEC Electropedia: available at <https://www.electropedia.org/>

¹ Impacted by EN ISO 139:2005/A1:2011

² Impacted by EN ISO 11737-1:2018/A1:2021

3.1
colony forming unit
CFU

unit by which the culturable number of microorganisms is expressed

Note 1 to entry: The culturable number is the number of microorganisms, single cells or aggregates, able to form colonies on a solid nutrient medium.

3.2
clean air suit

suit, used as working garment, intended and shown to minimize contamination of the operating room air from skin scales originating on the skin of persons wearing it

Note 1 to entry: A scrub suit is a working garment for operating room staff that does not need to meet the requirements for a clean air suit. The scrub suit is not primarily intended to prevent airborne dispersal from staff and can be designed and processed as the manufacturer thinks fit.

Note 2 to entry: A clean air suit consists of a coverall, or a blouse and a pair of trousers and can also include a barrier hood.

3.3
cleanliness

freedom from unwanted foreign matter

Note 1 to entry: Such matter can be microorganisms, organic residues or particulate matter.

3.3.1
cleanliness — microbial

freedom from population of viable micro-organisms on a product and/or a package

Note 1 to entry: In practical use, microbial cleanliness is often referred to as 'bioburden'.

3.4
infective agent

microorganism that has been shown to cause wound infections or that might cause infection in a member of the surgical team or the patient

3.5
manufacturer

natural or legal person with responsibility for the design, manufacture, packaging and labelling of a device before it is placed on the market under their own name, regardless of whether these operations are carried out by that person themselves or on their behalf by a third party

Note 1 to entry: For more details refer to the Medical Device Regulation (EU) 2017/745.

3.6
particle release

release of fibre fragments and other particles during mechanical stress simulating handling and use