

INTERNATIONAL STANDARD



This extended version of IEC 61010-2-030:2023 includes the content of the references made to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016

GROUP SAFETY PUBLICATION

**Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 2-030: Particular requirements for equipment having testing or measuring circuits**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2023 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 300 terminological entries in English and French, with equivalent terms in 19 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.



IEC 61010-2-030

Edition 3.0 2023-09
EXTENDED VERSION

INTERNATIONAL STANDARD



This extended version of IEC 61010-2-030:2023 includes the content of the references made to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016

GROUP SAFETY PUBLICATION

**Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 2-030: Particular requirements for equipment having testing or measuring circuits**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 19.080, 71.040.10

ISBN 978-2-8322-7562-7

Warning! Make sure that you obtained this publication from an authorized distributor.

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 1: General requirements

INTERPRETATION SHEET 1

This interpretation sheet has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

The text of this interpretation sheet is based on the following documents:

| | |
|-------------|------------------|
| ISH | Report on voting |
| 66/497A/ISH | 66/505/RVD |

Full information on the voting for the approval of this interpretation sheet can be found in the report on voting indicated in the above table.

IEC 61010-1:2010 contains a requirement in 6.8.3.1 pertaining to voltage testers for type tests as follows:

“The generator shall be able to supply a power of at least 500 VA.”

This has given rise to the following questions:

How does one interpret the requirement for voltage testers in 6.8.3.1 of IEC 61010-1:2010? Specifically, this subclause requires that “The generator shall be able to supply a power of at least 500 VA.” Does this requirement apply throughout the rated output range of the voltage tester? What is meant by the word “generator”? Is the “generator” the power supply within the voltage tester, or the voltage tester output, or something else?

Interpretation:

“A voltage tester used for type tests must be able to deliver at least 500 VA at its full-rated output voltage. It does not necessarily need to deliver 500 VA if set for lower voltages.

For example, a voltage tester that can deliver 100 mA at any test output voltage up to 5 000 V (and a current corresponding to 500 VA above 5 000 V) would meet the requirement.

The requirements for voltage testers used for routine (production line) tests are included in Annex F. The requirements of 6.8.3.1 do not apply to these voltage testers.”

CONTENTS

| | |
|--|-----------|
| FOREWORD..... | 10 |
| INTRODUCTION to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016..... | 14 |
| INTRODUCTION..... | 15 |
| 1 Scope and object..... | 16 |
| 1.1 Scope..... | 16 |
| 1.1.1 Equipment included in scope | 16 |
| 1.1.2 Equipment excluded from scope | 16 |
| 1.1.3 Computing equipment..... | 17 |
| 1.2 Object | 17 |
| 1.2.1 Aspects included in scope | 17 |
| 1.2.2 Aspects excluded from scope | 17 |
| 1.3 Verification | 17 |
| 1.4 Environmental conditions | 18 |
| 1.4.1 Normal environmental conditions | 18 |
| 1.4.2 Extended environmental conditions | 18 |
| 2 Normative references | 18 |
| 3 Terms and definitions | 21 |
| 3.1 Equipment and states of equipment..... | 21 |
| 3.2 Parts and accessories | 21 |
| 3.3 Quantities..... | 22 |
| 3.4 Tests | 22 |
| 3.5 Safety terms | 23 |
| 3.6 Insulation | 24 |
| 4 Tests | 26 |
| 4.1 General | 26 |
| 4.2 Sequence of tests..... | 26 |
| 4.3 Reference test conditions | 26 |
| 4.3.1 Environmental conditions..... | 26 |
| 4.3.2 State of equipment | 27 |
| 4.4 Testing in SINGLE FAULT CONDITION | 28 |
| 4.4.1 General | 28 |
| 4.4.2 Application of fault conditions | 29 |
| 4.4.3 Duration of tests | 31 |
| 4.4.4 Conformity after application of fault conditions..... | 31 |
| 5 Marking and documentation..... | 32 |
| 5.1 Marking | 32 |
| 5.1.1 General | 32 |
| 5.1.2 Identification..... | 32 |
| 5.1.3 MAINS supply | 33 |
| 5.1.4 Fuses | 34 |
| 5.1.5 TERMINALS, connections and operating devices..... | 35 |
| 5.1.6 Switches and circuit-breakers | 37 |
| 5.1.7 Equipment protected by DOUBLE INSULATION or REINFORCED INSULATION | 37 |
| 5.1.8 Field-wiring TERMINAL boxes | 37 |
| 5.2 Warning markings | 37 |

| | | |
|---------|--|----|
| 5.3 | Durability of markings..... | 38 |
| 5.4 | Documentation | 38 |
| 5.4.1 | General | 38 |
| 5.4.2 | Equipment RATINGS | 39 |
| 5.4.3 | Equipment installation | 40 |
| 5.4.4 | Equipment operation..... | 40 |
| 5.4.5 | Equipment maintenance and service | 41 |
| 5.4.6 | Integration into systems or effects resulting from special conditions | 41 |
| 6 | Protection against electric shock | 41 |
| 6.1 | General | 41 |
| 6.1.1 | Requirements | 41 |
| 6.1.2 | Exceptions..... | 42 |
| 6.2 | Determination of ACCESSIBLE parts..... | 42 |
| 6.2.1 | General | 42 |
| 6.2.2 | Examination..... | 42 |
| 6.2.3 | Openings above parts that are HAZARDOUS LIVE..... | 43 |
| 6.2.4 | Openings for pre-set controls..... | 43 |
| 6.3 | Limit values for ACCESSIBLE parts..... | 43 |
| 6.3.1 | Levels in NORMAL CONDITION..... | 43 |
| 6.3.2 | Levels in SINGLE FAULT CONDITION..... | 44 |
| 6.4 | Primary means of protection | 47 |
| 6.4.1 | General | 47 |
| 6.4.2 | ENCLOSURES and PROTECTIVE BARRIERS..... | 47 |
| 6.4.3 | BASIC INSULATION | 47 |
| 6.4.4 | Impedance..... | 47 |
| 6.5 | Additional means of protection in case of SINGLE FAULT CONDITIONS | 47 |
| 6.5.1 | General | 47 |
| 6.5.2 | PROTECTIVE BONDING | 48 |
| 6.5.3 | SUPPLEMENTARY INSULATION and REINFORCED INSULATION | 52 |
| 6.5.4 | PROTECTIVE IMPEDANCE | 53 |
| 6.5.5 | Automatic disconnection of the supply | 53 |
| 6.5.6 | Current- or voltage-limiting device | 53 |
| 6.6 | Connections to external circuits..... | 54 |
| 6.6.1 | General | 54 |
| 6.6.2 | TERMINALS for external circuits..... | 54 |
| 6.6.3 | Circuits with TERMINALS which are HAZARDOUS LIVE..... | 54 |
| 6.6.4 | TERMINALS for stranded conductors | 54 |
| 6.6.101 | Measuring circuit TERMINALS..... | 55 |
| 6.7 | Insulation requirements | 57 |
| 6.7.1 | The nature of insulation | 57 |
| 6.7.2 | Insulation for MAINS CIRCUITS of OVERVOLTAGE CATEGORY II with a nominal supply voltage up to 300 V | 60 |
| 6.7.3 | Insulation for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V | 63 |
| 6.8 | Procedure for voltage tests..... | 69 |
| 6.8.1 | General | 69 |
| 6.8.2 | Humidity preconditioning | 70 |
| 6.8.3 | Test procedures | 71 |
| 6.9 | Constructional requirements for protection against electric shock..... | 72 |

| | | |
|--------|--|----|
| 6.9.1 | General | 72 |
| 6.9.2 | Insulating materials | 72 |
| 6.9.3 | Colour coding | 72 |
| 6.10 | Connection to the MAINS supply source and connections between parts of equipment | 72 |
| 6.10.1 | MAINS supply cords | 72 |
| 6.10.2 | Fitting of non-detachable MAINS supply cords | 73 |
| 6.10.3 | Plugs and connectors | 75 |
| 6.11 | Disconnection from supply source | 75 |
| 6.11.1 | General | 75 |
| 6.11.2 | Exceptions | 75 |
| 6.11.3 | Requirements according to type of equipment | 76 |
| 6.11.4 | Disconnecting devices | 76 |
| 7 | Protection against mechanical HAZARDS | 77 |
| 7.1 | General | 77 |
| 7.2 | Sharp edges | 77 |
| 7.3 | Moving parts | 77 |
| 7.3.1 | General | 77 |
| 7.3.2 | Exceptions | 78 |
| 7.3.3 | RISK assessment for mechanical HAZARDS to body parts | 78 |
| 7.3.4 | Limitation of force and pressure | 79 |
| 7.3.5 | Gap limitations between moving parts | 80 |
| 7.4 | Stability | 82 |
| 7.5 | Provisions for lifting and carrying | 83 |
| 7.5.1 | General | 83 |
| 7.5.2 | Handles and grips | 83 |
| 7.5.3 | Lifting devices and supporting parts | 83 |
| 7.6 | Wall mounting | 83 |
| 7.7 | Expelled parts | 84 |
| 8 | Resistance to mechanical stresses | 84 |
| 8.1 | General | 84 |
| 8.2 | ENCLOSURE rigidity tests | 85 |
| 8.2.1 | Static test | 85 |
| 8.2.2 | Impact test | 85 |
| 8.3 | Drop test | 86 |
| 8.3.1 | Equipment other than HAND-HELD EQUIPMENT and DIRECT PLUG-IN EQUIPMENT | 86 |
| 8.3.2 | HAND-HELD EQUIPMENT and DIRECT PLUG-IN EQUIPMENT | 87 |
| 9 | Protection against the spread of fire and arc flash | 87 |
| 9.1 | General | 87 |
| 9.2 | Eliminating or reducing the sources of ignition within the equipment | 89 |
| 9.3 | Containment of fire within the equipment, should it occur | 89 |
| 9.3.1 | General | 89 |
| 9.3.2 | Constructional requirements | 89 |
| 9.4 | Limited-energy circuit | 92 |
| 9.5 | Requirements for equipment containing or using flammable liquids | 93 |
| 9.6 | Overcurrent protection | 93 |
| 9.6.1 | General | 93 |
| 9.6.2 | PERMANENTLY CONNECTED EQUIPMENT | 94 |

| | | |
|---------|--|-----|
| 9.6.3 | Other equipment..... | 94 |
| 9.101 | Protection of measuring circuits | 94 |
| 9.101.1 | General | 94 |
| 9.101.2 | Protection against mismatches of inputs and ranges..... | 94 |
| 9.101.3 | Protection against MAINS overvoltages | 97 |
| 10 | Equipment temperature limits and resistance to heat..... | 98 |
| 10.1 | Surface temperature limits for protection against burns | 98 |
| 10.2 | Temperatures of windings | 99 |
| 10.3 | Other temperature measurements | 99 |
| 10.4 | Conduct of temperature tests | 100 |
| 10.4.1 | General | 100 |
| 10.4.2 | Temperature measurement of heating equipment | 100 |
| 10.4.3 | Equipment intended for installation in a cabinet or a wall..... | 100 |
| 10.5 | Resistance to heat..... | 101 |
| 10.5.1 | Integrity of CLEARANCES and CREEPAGE DISTANCES..... | 101 |
| 10.5.2 | Non-metallic ENCLOSURES | 101 |
| 10.5.3 | Insulating material | 101 |
| 11 | Protection against HAZARDS from fluids and solid foreign objects | 102 |
| 11.1 | General | 102 |
| 11.2 | Cleaning..... | 102 |
| 11.3 | Spillage..... | 103 |
| 11.4 | Overflow..... | 103 |
| 11.5 | Battery electrolyte | 103 |
| 11.6 | Equipment RATED with a degree of ingress protection (IP code)..... | 103 |
| 11.6.1 | General | 103 |
| 11.6.2 | Conditions for testing..... | 104 |
| 11.6.3 | Protection against solid foreign objects (including dust)..... | 104 |
| 11.6.4 | Protection against water | 104 |
| 11.7 | Fluid pressure and leakage | 105 |
| 11.7.1 | Maximum pressure | 105 |
| 11.7.2 | Leakage and rupture at high pressure | 105 |
| 11.7.3 | Leakage from low-pressure parts..... | 106 |
| 11.7.4 | Overpressure safety device | 106 |
| 12 | Protection against radiation, including laser sources, and against sonic and ultrasonic pressure | 107 |
| 12.1 | General..... | 107 |
| 12.2 | Equipment producing ionizing radiation | 107 |
| 12.2.1 | Ionizing radiation | 107 |
| 12.2.2 | Accelerated electrons | 108 |
| 12.3 | Optical radiation | 108 |
| 12.4 | Microwave radiation | 109 |
| 12.5 | Sonic and ultrasonic pressure | 109 |
| 12.5.1 | Sound level | 109 |
| 12.5.2 | Ultrasonic pressure..... | 110 |
| 12.6 | Laser sources..... | 110 |
| 13 | Protection against liberated gases and substances, explosion and implosion | 111 |
| 13.1 | Poisonous and injurious gases and substances..... | 111 |
| 13.2 | Explosion and implosion | 111 |
| 13.2.1 | Components | 111 |

| | | |
|------------------------|---|-----|
| 13.2.2 | Batteries and battery charging | 111 |
| 13.2.3 | Implosion of cathode ray tubes | 112 |
| 14 | Components and subassemblies | 112 |
| 14.1 | General | 112 |
| 14.2 | Motors | 113 |
| 14.2.1 | Motor temperatures | 113 |
| 14.2.2 | Series excitation motors | 113 |
| 14.3 | Overtemperature protection devices | 114 |
| 14.4 | Fuse holders | 114 |
| 14.5 | MAINS voltage selection devices | 114 |
| 14.6 | MAINS transformers tested outside equipment..... | 114 |
| 14.7 | Printed wiring boards..... | 115 |
| 14.8 | Circuits used to limit TRANSIENT OVERVOLTAGES..... | 115 |
| 14.101 | Probe assemblies and accessories | 115 |
| 15 | Protection by interlocks | 116 |
| 15.1 | General | 116 |
| 15.2 | Prevention of reactivating | 116 |
| 15.3 | Reliability | 116 |
| 16 | HAZARDS resulting from application | 116 |
| 16.1 | REASONABLY FORESEEABLE MISUSE..... | 116 |
| 16.2 | Ergonomic aspects | 116 |
| 17 | RISK assessment | 117 |
| 101 | Measuring circuits | 117 |
| 101.1 | General | 117 |
| 101.2 | Current measuring circuits..... | 118 |
| 101.3 | Indicating devices..... | 118 |
| 101.3.1 | General | 118 |
| 101.3.2 | Battery level | 118 |
| 101.3.3 | Over-range indication | 119 |
| 101.3.4 | Permanent overvoltages | 119 |
| Annex A (normative) | Measuring circuits for touch current (see 6.3)..... | 121 |
| Annex B (normative) | Standard test fingers (see 6.2) | 124 |
| Annex C (normative) | Measurement of CLEARANCES and CREEPAGE DISTANCES | 126 |
| Annex D (normative) | Parts between which insulation requirements are specified (see 6.4 and 6.5.3) | 130 |
| Annex E (informative) | Guideline for reduction of POLLUTION DEGREES | 133 |
| Annex F (normative) | ROUTINE TESTS..... | 134 |
| Annex G (informative) | Leakage and rupture from fluids under pressure | 136 |
| Annex H (normative) | Qualification of conformal coatings for protection against POLLUTION | 141 |
| Annex I (informative) | Line-to-neutral voltages for common MAINS supply systems | 144 |
| Annex J (informative) | RISK assessment | 146 |
| Annex K (normative) | Insulation requirements not covered by 6.7 | 149 |
| Annex L (informative) | Index of defined terms | 176 |
| Annex AA (normative) | MEASUREMENT CATEGORIES | 178 |
| Annex BB (informative) | HAZARDS pertaining to measurements performed in certain environments | 181 |

| | |
|---|-----|
| Annex CC (informative) 4 mm "banana" TERMINALS | 184 |
| Annex DD (informative) Flowchart for insulation according to the type of circuit | 186 |
| Annex EE (informative) Determination of CLEARANCES for Table 101 | 189 |
| Bibliography..... | 190 |
| Figure 1 – Measurements through openings in ENCLOSURES..... | 43 |
| Figure 2 – Maximum duration of short-term ACCESSIBLE voltages in SINGLE FAULT CONDITION (see 6.3.2 a)) | 45 |
| Figure 3 – Capacitance level versus voltage in NORMAL CONDITION and SINGLE FAULT CONDITION (see 6.3.1 c) and 6.3.2 c)) | 46 |
| Figure 4 – Acceptable arrangement of protective means against electric shock | 48 |
| Figure 5 – Examples of binding screw assemblies | 50 |
| Figure 101 – Duration of current flow versus body current for a.c. and d.c. currents..... | 52 |
| Figure 6 – Distance between conductors on an interface between two layers..... | 61 |
| Figure 7 – Distance between adjacent conductors along an interface of two inner layers..... | 62 |
| Figure 8 – Distance between adjacent conductors located between the same two layers..... | 63 |
| Figure 9 – Detachable MAINS supply cords and connections | 73 |
| Figure 10 – Impact test using a sphere | 86 |
| Figure 11 – Flow chart to explain the requirements for protection against the spread of fire | 88 |
| Figure 12 – Baffle | 91 |
| Figure 13 – Area of the bottom of an ENCLOSURE to be constructed as specified in 9.3.2 c) 1) | 91 |
| Figure 14 – Ball-pressure test apparatus..... | 102 |
| Figure 16 – Ratio between test pressure and maximum working pressure..... | 106 |
| Figure 15 – Flow chart for conformity options 14.1 a), b), c) and d)..... | 113 |
| Figure A.1 – Measuring circuit for a.c. with frequencies up to 1 MHz and for d.c. | 121 |
| Figure A.2 – Measuring circuits for sinusoidal a.c. with frequencies up to 100 Hz and for d.c. | 122 |
| Figure A.3 – Current measuring circuit for electrical burns | 123 |
| Figure A.4 – Current measuring circuit for wet contact | 123 |
| Figure B.1 – Rigid test finger | 124 |
| Figure B.2 – Jointed test finger | 125 |
| Figure C.1 – Examples of methods of measuring CLEARANCES and CREEPAGE DISTANCES..... | 129 |
| Figures D.1a) to d) – Protection between HAZARDOUS LIVE circuits and ACCESSIBLE parts..... | 131 |
| Figures D.1e) to h) – Protection between HAZARDOUS LIVE circuits and circuits with ACCESSIBLE external TERMINALS | 131 |
| Figures D.2 a) and D.2 b) – Protection between a HAZARDOUS LIVE internal circuit and an ACCESSIBLE part which is not bonded to other ACCESSIBLE parts..... | 132 |
| Figures D.2 c) and D.2 d) – Protection between a HAZARDOUS LIVE primary circuit and circuits which have ACCESSIBLE external TERMINALS | 132 |
| Figure D.3 – Protection of external ACCESSIBLE TERMINALS of two HAZARDOUS LIVE circuits..... | 132 |
| Figure G.1 – Conformity verification process (see G.2) | 137 |

| | |
|---|-----|
| Figure H.1 – Test sequence and conformity | 143 |
| Figure J.1 – Iterative process of RISK assessment and RISK reduction | 146 |
| Figure J.2 – RISK reduction | 147 |
| Figure K.1 – Distance between conductors on an interface between two layers | 153 |
| Figure K.2 – Distance between adjacent conductors along an interface of an inner layer | 154 |
| Figure K.3 – Distance between adjacent conductors located between the same two layers..... | 155 |
| Figure K.4 – Example of recurring peak voltage | 167 |
| Figure K.101 – Test circuit for evaluation of TRANSIENT OVERVOLTAGE attenuation..... | 170 |
| Figure AA.1 – Example to identify the locations of MEASUREMENT CATEGORIES | 179 |
| Figure CC.1 – Recommended dimensions of 4 mm TERMINALS | 185 |
| Figure DD.1 – Requirements for CLEARANCE, CREEPAGE DISTANCE and solid insulation..... | 188 |
| Table 1 – Symbols | 34 |
| Table 2 – Tightening torque for binding screw assemblies | 50 |
| Table 101 – CLEARANCES for unmated measuring circuit TERMINALS..... | 56 |
| Table 3 – Multiplication factors for CLEARANCES of equipment RATED for operation at altitudes up to 5 000 m | 58 |
| TABLE 4 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V | 60 |
| Table 5 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V..... | 61 |
| Table 6 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY II up to 300 V..... | 65 |
| Table 7 – CREEPAGE DISTANCES for secondary circuits | 66 |
| Table 8 – Minimum values for distance or thickness (see 6.7.3.4.2 to 6.7.3.4.4) | 68 |
| Table 9 – Distances between TERMINALS and foil..... | 69 |
| Table 10 – Correction factors according to test site altitude for test voltages for CLEARANCES | 70 |
| Table 11 – Values for physical tests on cord anchorages | 74 |
| Table 12 – Protective measures against mechanical HAZARDS to body parts..... | 79 |
| Table 13 – Minimum maintained gaps to prevent crushing for different body parts | 81 |
| Table 14 – Maximum gaps to prevent access for different body parts..... | 81 |
| Table 15 – Impact energy levels, test height and corresponding IK codes | 86 |
| Table 16 – Acceptable perforation of the bottom of an ENCLOSURE | 90 |
| Table 17 – Limits of maximum available current..... | 92 |
| Table 18 – Values for overcurrent protection devices | 93 |
| Table 19 – Surface temperature limits in NORMAL CONDITION..... | 99 |
| Table 20 – Maximum temperatures for insulation material of windings | 99 |
| Table 22 – Lamp or lamp systems considered photobiologically safe | 109 |
| Table 23 – Lamp or lamp systems considered photobiologically safe under certain conditions | 109 |
| Table 21 – Impulse voltages for OVERVOLTAGECATEGORY II..... | 115 |
| Table C.1 – Dimensions of X..... | 126 |

| | |
|---|-----|
| Table E.1 – Environmental situations | 133 |
| Table E.2 – Reduction of POLLUTION DEGREES | 133 |
| Table F.1 – Test voltages for ROUTINE TESTS of MAINS CIRCUITS | 135 |
| Table G.1 – Test pressures for equipment with pressures above 14 Mpa | 139 |
| Table H.1 – Test parameters, test conditions and test procedures | 142 |
| Table I.1 – Line-to-neutral voltages for common MAINS supply systems | 144 |
| Table J.1 – Severity of harm | 148 |
| Table J.2 – Probability of harm | 148 |
| Table J.3 – Risk category | 148 |
| Table K.1 – Multiplication factors for CLEARANCES for equipment RATED for operation at altitudes up to 5 000 m | 150 |
| Table K.2 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY II above 300 V | 150 |
| Table K.3 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY III | 151 |
| Table K.4 – CLEARANCES and CREEPAGE DISTANCES for MAINS CIRCUITS of OVERVOLTAGE CATEGORY IV | 151 |
| Table K.5 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY II above 300 V | 152 |
| Table K.6 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY III | 152 |
| Table K.7 – Test voltages for solid insulation in MAINS CIRCUITS of OVERVOLTAGE CATEGORY IV | 152 |
| Table K.8 – Test voltages for testing long-term stress of solid insulation in MAINS CIRCUITS | 153 |
| Table K.9 – Minimum values for distance or thickness of solid insulation | 154 |
| Table K.10 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY II above 300 V | 157 |
| Table K.11 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY III | 158 |
| Table K.12 – CLEARANCES and test voltages for secondary circuits derived from MAINS CIRCUITS of OVERVOLTAGE CATEGORY IV | 159 |
| Table K.13 – CREEPAGE DISTANCES for secondary circuits | 160 |
| Table K.14 – Minimum values for distance or thickness (see K.2.4.2 to K.2.4.4) | 162 |
| Table K.15 – CLEARANCE values for the calculation of K.3.2 | 165 |
| Table K.16 – Test voltages based on CLEARANCES | 165 |
| Table K.17 – CLEARANCES for BASIC INSULATION in circuits having recurring peak voltages or WORKING VOLTAGES with frequencies above 30 kHz | 168 |
| Table K.101 – Impulse voltages for circuits connected to MAINS | 170 |
| Table K.102 – CLEARANCES for measuring circuits RATED for MEASUREMENT CATEGORIES | 171 |
| Table K.103 – Impulse test voltages for testing electric strength of solid insulation for measuring circuits RATED for MEASUREMENT CATEGORIES | 172 |
| Table K.104 – a.c. test voltages for testing electric strength of solid insulation for measuring circuits RATED for MEASUREMENT CATEGORIES | 173 |
| Table K.105 – Minimum values for distance or thickness of solid insulation for measuring circuits RATED for MEASUREMENT CATEGORIES | 174 |
| Table AA.1 – Characteristics of MEASUREMENT CATEGORIES | 180 |
| Table EE.1 – CLEARANCE values for Table 101 | 189 |

INTERNATIONAL ELECTROTECHNICAL COMMISSION

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 1: General requirements

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

This extended version (EXV) of the official IEC Standard provides the user with the comprehensive content of the Standard.

IEC 61010-2-030:2023 EXV includes the content of IEC 61010-2-030:2023, and the references made to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016.

The specific content of IEC 61010-2-030:2023 is displayed on a blue background.

IEC 61010-2-030 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment. It is an International Standard.

It has the status of a group safety publication in accordance with IEC Guide 104.

This third edition cancels and replaces the second edition published in 2017. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) in 1.2.1, requirements for protection against HAZARDS which could occur from reading a voltage have been added to the scope;
- b) Clause 2, all normative references have been dated and new normative references have been added;
- c) in 4.3.2.5, requirements for power supply have been modified;
- d) in 4.3.2.6, requirements for inputs/outputs have been modified;
- e) in 4.4.2.101, a new subclause about surge protective devices has been added;
- f) in 5.1.5.101.2, minimum RATINGS for voltage of measuring TERMINALS are required;
- g) Subclause 6.6.101 modifies 6.6.101 and 6.6.102 of previous edition:
 - 1) in 6.6.101.1, insulating material of group I may be allowed for determination of CREEPAGE DISTANCES of measuring circuit TERMINALS;
 - 2) In 6.6.101.2, CLEARANCES and CREEPAGE DISTANCES up to 3 000 V for measuring circuit TERMINALS in unmated position have been defined;
 - 3) in 6.6.101.3, requirements for measuring circuit TERMINALS in partially mated position have been specified;
 - 4) in 6.6.101.4, requirements for measuring circuit TERMINALS in mated position have been specified;
 - 5) Subclause 6.6.101.5 replaces 6.6.102;
- h) Subclause 9.101 to consider the protection of measuring circuits against the spread of fire and arc flash has been added and Table 102 has been replaced by Table K.101;
- i) in 9.101.2, relocation of 101.3 of previous edition;
- j) in 9.101.3, relocation of 101.4 of previous edition, extension to MEASUREMENT CATEGORY II and reference to IEC 61000-4-5 for tests;
- k) in 14.101, relocation of 14.102 and 14.101 of previous edition has been removed;
- l) in 101.3, relocation of 101.5 of previous edition, and more requirements added against HAZARD occurring from reading a voltage value;
- m) in K.2.1, another method for determination of CLEARANCES of secondary circuits is proposed;
- n) in K.3.2, new Table K.15 and Table K.16 for CLEARANCE calculation;
- o) in K.101.4.1, new Table K.103 and Table K.104 replace Table K.102, Table K.103 and Table K.104;
- p) in K.101.4, the subclause has been reviewed; Tables and tests for solid insulation have been modified; Table K.105 replaces Table K.9;
- q) Table K.101, replacement of Table K.106;
- r) Clause K.4, redraft of the clause to propose a method for determination of U_t for circuits which reduce TRANSIENT OVERVOLTAGES;
- s) Annex AA: Figure AA.1 has been redesigned;
- t) Annex EE: addition of a new informative annex for determination of CLEARANCES for Table 101.

The text of this International Standard is based on the following documents:

| | |
|-------------|------------------|
| Draft | Report on voting |
| 66/786/FDIS | 66/796/RVD |

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

A list of all parts of the IEC 61010 series, under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

This document is to be used in conjunction with IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016.

This document supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for equipment having testing or measuring circuits*.

Where a particular subclause of IEC 61010-1 is not mentioned in this document, that subclause applies as far as is reasonable. Where this document states "addition", "modification", "replacement", or "deletion", the relevant requirement, test specification or note in IEC 61010-1 should be adapted accordingly.

In this standard:

- the following print types are used:
 - requirements: in roman type;
 - NOTES: in small roman type;
 - *conformity and tests: in italic type*;
 - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS;
- subclauses, figures, tables and notes which are additional to those in IEC 61010-1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION to IEC 61010-1:2010 and IEC 61010-1:2010/AMD1:2016

This International Standard specifies the safety requirements that are generally applicable to all equipment within its scope. For certain types of equipment, these requirements will be supplemented or modified by the special requirements of one, or more than one, particular part 2 of the standard which must be read in conjunction with the part 1 requirements.

INTRODUCTION

IEC 61010-1 specifies the safety requirements that are generally applicable to all equipment within its scope. For certain types of equipment, the requirements of IEC 61010-1 and its amendment will be supplemented or modified by the special requirements of one or more standard from the IEC 61010-2 series which is/are read in conjunction with the requirements of IEC 61010-1.

- 1) This document specifies the safety requirements for equipment with testing or measuring circuits which are connected for test or measurement purposes to devices or circuits outside the measurement equipment itself.
- 2) IEC 61010-2-032:2023 specifies the safety requirements for hand-held and hand-manipulated current sensors for measuring, detecting, injecting current, or indicating current waveforms on circuits without physically opening the current path of the circuit being measured.

Most of the requirements of this document have been included in IEC 61010-2-032:2023. Equipment within the scopes of both this document and IEC 61010-2-032:2023 is considered to be covered by the requirements of IEC 61010-2-032:2023.

However, for current sensors in combined equipment with protective bonding and automatic disconnection of the supply, this document and IEC 61010-2-032:2023 are read in conjunction.

- 3) IEC 61010-2-033:2023 specifies the safety requirements for hand-held multimeters and other meters for domestic and professional use, capable of measuring mains voltage, intended to measure voltage and other electrical quantities such as resistance or current.

All relevant requirements of this document have been included in IEC 61010-2-033:2023.

- 4) IEC 61010-2-034:2023 specifies the safety requirements for measurement equipment for insulation resistance and test equipment for electric strength which are connected to units, lines or circuits for test or measurement purposes.

All relevant requirements of this document have been included in IEC 61010-2-034:2023. However, for equipment within the scope of IEC 61010-2-032:2023 and IEC 61010-2-034:2023, these standards are read in conjunction.

IEC 61010-031 specifies the safety requirements for hand-held and hand-manipulated probe assemblies and their related accessories intended to be used in particular with equipment in the scope of this document, IEC 61010-2-032, IEC 61010-2-033 and IEC 61010-2-034. These probe assemblies are for non-contact or direct electrical connection between a part and electrical test and measurement equipment. They may be fixed to the equipment or be detachable accessories for the equipment.