



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN ISO 10555-1:2023

Intravascular catheters - Sterile and single-use catheters - Part 1: General requirements (ISO 10555-1:2023)

Cathéters intravasculaires - Cathéters
stériles et non réutilisables - Partie 1:
Exigences générales (ISO 10555-1:2023)

Intravaskuläre Katheter - Sterile Katheter
zur einmaligen Verwendung - Teil 1:
Allgemeine Anforderungen (ISO
10555-1:2023)



National Foreword

This European Standard EN ISO 10555-1:2023 was adopted as Luxembourgish Standard ILNAS-EN ISO 10555-1:2023.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

ILNAS-EN ISO 10555-1:2023

EUROPEAN STANDARD **EN ISO 10555-1**

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2023

ICS 11.040.25

Supersedes EN ISO 10555-1:2013

English Version

**Intravascular catheters - Sterile and single-use catheters -
Part 1: General requirements (ISO 10555-1:2023)**

Cathéters intravasculaires - Cathéters stériles et non
réutilisables - Partie 1: Exigences générales (ISO
10555-1:2023)

Intravaskuläre Katheter - Sterile Katheter zur
einmaligen Verwendung - Teil 1: Allgemeine
Anforderungen (ISO 10555-1:2023)

This European Standard was approved by CEN on 24 November 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

European foreword

This document (EN ISO 10555-1:2023) has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2024, and conflicting national standards shall be withdrawn at the latest by May 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10555-1:2013.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10555-1:2023 has been approved by CEN as EN ISO 10555-1:2023 without any modification.

Intravascular catheters — Sterile and single-use catheters —

Part 1: General requirements

*Cathéters intravasculaires — Cathéters stériles et non réutilisables —
Partie 1: Exigences générales*



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	5
4.1 Risk approach.....	5
4.2 Usability engineering.....	5
4.3 Sterilization.....	5
4.4 Shelf life.....	5
4.5 Detectability.....	5
4.6 Biocompatibility.....	5
4.7 Surface.....	5
4.8 Corrosion resistance.....	6
4.9 Peak tensile force.....	6
4.10 Freedom from leakage during pressurization.....	7
4.11 Freedom from leakage during aspiration.....	7
4.12 Hubs.....	7
4.13 Flowrate.....	8
4.14 Power injection burst pressure.....	8
4.15 Packaging system.....	8
4.16 Simulated use, kink and/or torque testing to consider depending on device design, intended use, and risk analysis.....	8
4.17 Coating integrity and/or particulate testing to consider depending on device design, intended use, and risk analysis.....	9
4.18 Distal tip stiffness testing to consider for neurovascular applications.....	9
5 Designation of nominal size	9
5.1 Nominal outside diameter.....	9
5.2 Nominal inside diameter.....	9
5.3 Nominal effective length.....	9
6 Information to be supplied with the catheter	10
6.1 General.....	10
6.2 Marking on the device and/or primary packaging.....	10
6.3 Instructions for use.....	11
6.4 Marking on the secondary packaging.....	11
Annex A (normative) Test method for corrosion resistance	12
Annex B (normative) Method for determining peak tensile force	13
Annex C (normative) Test method for liquid leakage under pressure	16
Annex D (normative) Test method for air leakage into hub assembly during aspiration	18
Annex E (normative) Determination of flowrate through catheter	20
Annex F (normative) Test for burst pressure under static conditions	22
Annex G (normative) Power injection tests for flowrate and device pressure (only for products indicated for power injection)	25
Annex H (informative) Units of measurement systems other than those specified in this document	30
Annex I (normative) Test method for air leakage under water	32
Annex J (informative) Rationale and guidance	34
Annex K (informative) Test methods for distal tip stiffness for neurovascular applications	41