

ILNAS

Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN ISO 2612:2023

Analysis of natural gas - Biomethane - Determination of ammonia content by Tunable Diode Laser Absorption Spectroscopy (ISO 2612:2023)

Analyse du gaz naturel - Biométhane -
Détermination de la teneur en ammoniac
par spectroscopie d'absorption laser à
diode accordable (ISO 2612:2023)

Analyse von Erdgas - Biomethan -
Bestimmung von Ammoniakanteil durch
Absorptionsspektroskopie mittels
durchstimmbarer Laserdioden (ISO

National Foreword

This European Standard EN ISO 2612:2023 was adopted as Luxembourgish Standard ILNAS-EN ISO 2612:2023.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

ILNAS-EN ISO 2612:2023
EUROPEAN STANDARD **EN ISO 2612**
NORME EUROPÉENNE
EUROPÄISCHE NORM

December 2023

ICS 75.060

English Version

Analysis of natural gas - Biomethane - Determination of ammonia content by Tuneable Diode Laser Absorption Spectroscopy (ISO 2612:2023)

Analyse du gaz naturel - Biométhane - Détermination de la teneur en ammoniac par spectroscopie d'absorption laser à diode accordable (ISO 2612:2023)

Analyse von Erdgas - Biomethan - Bestimmung von Ammoniakanteil durch Absorptionsspektroskopie mittels durchstimmbarer Laserdioden (ISO 2612:2023)

This European Standard was approved by CEN on 20 November 2023.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents

	Page
European foreword.....	3

European foreword

This document (EN ISO 2612:2023) has been prepared by Technical Committee ISO/TC 193 "Natural gas" in collaboration with Technical Committee CEN/TC 408 "Natural gas and biomethane for use in transport and biomethane for injection in the natural gas grid" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2024, and conflicting national standards shall be withdrawn at the latest by June 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 2612:2023 has been approved by CEN as EN ISO 2612:2023 without any modification.

First edition
2023-12

**Analysis of natural gas — Biomethane
— Determination of ammonia content
by tuneable diode laser absorption
spectroscopy**

*Analyse du gaz naturel — Biométhane — Détermination de la teneur
en ammoniac par spectroscopie d'absorption laser à diode accordable*



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2023

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Safety precautions	3
5 Principle	3
6 Apparatus	3
6.1 NH ₃ analyser	3
6.2 Gas delivery system	3
6.3 Temperature sensor	4
6.4 Pressure sensor	4
6.5 Calibration equipment	4
6.5.1 General	4
6.5.2 Pressure regulators for the NH ₃ cylinders	4
6.5.3 Flow controller	4
6.5.4 Flow meter	4
6.5.5 Dilution device	4
6.5.6 Output manifold	5
7 Reagents and materials	5
7.1 Methane	5
7.2 Calibration gases	5
7.3 Inert gas	5
8 Sampling	5
8.1 General	5
8.2 Construction materials	5
8.3 Cleanliness	6
8.4 Installation of the calibration gas cylinder	6
8.5 Pressure control	6
8.6 Purging of reduction valve and transfer lines	6
8.7 Flow control	6
8.8 Leak control	7
9 Calibration	7
9.1 Calibration procedures	7
9.2 Frequency of calibration	7
9.2.1 Multipoint calibration	7
9.2.2 Zero and span point calibration	7
10 Interferences	7
10.1 Interfering absorbers	7
10.2 Matrix gas	7
10.3 Secondary level spectroscopic effects: Gas temperature, gas pressure, spatial homogeneity	8
10.4 Humidity and carbon dioxide	8
11 Measurement procedure	8
12 Expression of results	8
12.1 Quantities and units	8
12.2 Uncertainty	8
13 Test report	9