



Institut luxembourgeois de la normalisation
de l'accréditation, de la sécurité et qualité
des produits et services

ILNAS-EN ISO 17294-1:2024

Water quality - Application of inductively coupled plasma mass spectrometry (ICP-MS) - Part 1: General requirements (ISO

Qualité de l'eau - Application de la
spectrométrie de masse avec plasma à
couplage inductif (ICP- MS) - Partie 1:
Exigences générales (ISO 17294-1:2024)

Wasserbeschaffenheit - Anwendung der
induktiv gekoppelten Plasma-
Massenspektrometrie (ICP-MS) - Teil 1:
Allgemeine Anforderungen (ISO

04/2024



National Foreword

This European Standard EN ISO 17294-1:2024 was adopted as Luxembourgish Standard ILNAS-EN ISO 17294-1:2024.

Every interested party, which is member of an organization based in Luxembourg, can participate for FREE in the development of Luxembourgish (ILNAS), European (CEN, CENELEC) and International (ISO, IEC) standards:

- Participate in the design of standards
- Foresee future developments
- Participate in technical committee meetings

<https://portail-qualite.public.lu/fr/normes-normalisation/participer-normalisation.html>

THIS PUBLICATION IS COPYRIGHT PROTECTED

Nothing from this publication may be reproduced or utilized in any form or by any mean - electronic, mechanical, photocopying or any other data carries without prior permission!

ILNAS-EN ISO 17294-1:2024

EUROPEAN STANDARD **EN ISO 17294-1**

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2024

ICS 13.060.50

Supersedes EN ISO 17294-1:2006

English Version

**Water quality - Application of inductively coupled plasma
mass spectrometry (ICP-MS) - Part 1: General
requirements (ISO 17294-1:2024)**

Qualité de l'eau - Application de la spectrométrie de
masse avec plasma à couplage inductif (ICP- MS) -
Partie 1: Exigences générales (ISO 17294-1:2024)

Wasserbeschaffenheit - Anwendung der induktiv
gekoppelten Plasma-Massenspektrometrie (ICP-MS) -
Teil 1: Allgemeine Anforderungen (ISO 17294-1:2024)

This European Standard was approved by CEN on 20 January 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3

ILNAS-EN ISO 17294-1:2024 - Preview only Copy via ILNAS e-Shop

European foreword

This document (EN ISO 17294-1:2024) has been prepared by Technical Committee ISO/TC 147 "Water quality" in collaboration with Technical Committee CEN/TC 230 "Water analysis" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2024, and conflicting national standards shall be withdrawn at the latest by October 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 17294-1:2006.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 17294-1:2024 has been approved by CEN as EN ISO 17294-1:2024 without any modification.



**Water quality — Application of
inductively coupled plasma mass
spectrometry (ICP-MS) —**

**Part 1:
General requirements**

*Qualité de l'eau — Application de la spectrométrie de masse avec
plasma à couplage inductif (ICP-MS) —*

Partie 1: Exigences générales

**Second edition
2024-03**



COPYRIGHT PROTECTED DOCUMENT

© ISO 2024

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Apparatus	4
5.1 General	4
5.2 Sample introduction	4
5.2.1 General	4
5.2.2 Pump	5
5.2.3 Nebulizer	5
5.2.4 Spray chamber	6
5.2.5 Other systems	6
5.3 Torch and plasma	6
5.4 Gas and gas control	7
5.5 Generator	7
5.6 Transfer of the ions to the mass spectrometer	7
5.7 Mass spectrometer	8
5.7.1 General	8
5.7.2 Lens system	8
5.7.3 Collision or reaction cell	8
5.7.4 Analyser	8
5.7.5 Detector	9
5.7.6 Alternative mass spectrometers and types of instruments	10
5.8 Signal processing and instrument control	11
6 Interferences by concomitant elements	11
6.1 General	11
6.2 Spectral interferences	11
6.2.1 General	11
6.2.2 Possible elimination strategies for polyatomic ion interferences	12
6.3 Non-spectral interferences	13
6.3.1 General	13
6.3.2 Interferences in the nebulization process	13
6.3.3 Interferences in the plasma	13
6.3.4 Interferences in the interface or lens area	14
6.3.5 Possible elimination strategies for non-spectral interferences (matrix effects)	14
7 Adjustment of the apparatus	17
7.1 General	17
7.2 Tuning the apparatus	17
7.2.1 General	17
7.2.2 Alignment of the plasma	18
7.2.3 Mass calibration	18
7.2.4 Resolution	18
7.2.5 Detector	18
7.3 Verification of instrument performance criteria	19
8 Preparatory steps	19
8.1 General	19
8.2 Choice of isotopes	20
8.3 Choice of instrumental settings	20
8.4 Choice of integration time	21
8.5 Choice of reference elements — Internal standards	22