

INTERNATIONAL STANDARD

IEC 60601-2-51

First edition
2003-02

Medical electrical equipment –

Part 2-51: Particular requirements for safety, including essential performance, of recording and analysing single channel and multichannel electrocardiographs

Appareils électromédicaux –

Partie 2-51: Règles particulières de sécurité et performances essentielles des électrocardiographes enregistreurs et analyseurs mono et multi-canaux



Reference number
IEC 60601-2-51:2003(E)

Publication numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series. For example, IEC 34-1 is now referred to as IEC 60034-1.

Consolidated editions

The IEC is now publishing consolidated versions of its publications. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Further information on IEC publications

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology. Information relating to this publication, including its validity, is available in the IEC Catalogue of publications (see below) in addition to new editions, amendments and corrigenda. Information on the subjects under consideration and work in progress undertaken by the technical committee which has prepared this publication, as well as the list of publications issued, is also available from the following:

- **IEC Web Site** (www.iec.ch)

- **Catalogue of IEC publications**

The on-line catalogue on the IEC web site (http://www.iec.ch/searchpub/cur_fut.htm) enables you to search by a variety of criteria including text searches, technical committees and date of publication. On-line information is also available on recently issued publications, withdrawn and replaced publications, as well as corrigenda.

- **IEC Just Published**

This summary of recently issued publications (http://www.iec.ch/online_news/justpub/jp_entry.htm) is also available by email. Please contact the Customer Service Centre (see below) for further information.

- **Customer Service Centre**

If you have any questions regarding this publication or need further assistance, please contact the Customer Service Centre:

Email: custserv@iec.ch
Tel: +41 22 919 02 11
Fax: +41 22 919 03 00

INTERNATIONAL STANDARD

IEC 60601-2-51

First edition
2003-02

Medical electrical equipment –

Part 2-51: Particular requirements for safety, including essential performance, of recording and analysing single channel and multichannel electrocardiographs

Appareils électromédicaux –

Partie 2-51: Règles particulières de sécurité et performances essentielles des électrocardiographes enregistreurs et analyseurs mono et multi-canaux

© IEC 2003 — Copyright - all rights reserved

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission, 3, rue de Varembé, PO Box 131, CH-1211 Geneva 20, Switzerland
Telephone: +41 22 919 02 11 Telefax: +41 22 919 03 00 E-mail: inmail@iec.ch Web: www.iec.ch



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE

XC

For price, see current catalogue

CONTENTS

FOREWORD	4
INTRODUCTION	6
SECTION ONE – GENERAL	
1 Scope and object	7
2 Terminology and definitions	8
4 General requirements for tests	11
6 Identification, marking and documents	12
SECTION TWO – ENVIRONMENTAL CONDITIONS	
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50 Accuracy of operating data	14
50.101 Automated measurements on ECGs (for ANALYSING ELECTROCARDIOGRAPHS)	14
50.102 Automated ECG interpretation (for ANALYSING ELECTROCARDIOGRAPHS)	19
51 Protection against hazardous output	23
51.101 LEADS	23
51.102 Input circuit	27
51.103 CALIBRATION	28
51.104 SENSITIVITY	29
51.105 Reduction of the effects of unwanted external voltages	29
51.106 Base-line	30
51.107 Distortion	32
51.108 Printing, electronic storage and transmission	34
51.109 Use with cardiac pacemakers	36
SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS; ENVIRONMENTAL TESTS	
SECTION TEN – CONSTRUCTIONAL REQUIREMENTS	
56 Components and general assembly	37
56.7 BATTERIES	37
Appendix L (normative) References – Publications mentioned in this standard	43
Annex AA (informative) General guidance and rationale	44
Annex BB (informative) ELECTRODES, their positions, identifications and colour codes	51
Annex CC (informative) LEADS and their identification (other than described in 51.101)	53
Annex DD (informative) Polarity of PATIENT LEADS (other than those specified in 51.101)	55
Annex EE (informative) Additional marking of electrodes	56
Annex FF (informative) Noise	58

Annex GG (normative) Definitions and rules for the measurement of ELECTROCARDIOGRAMS.....	60
Annex HH (normative) Calibration and test data sets	66
Annex II (informative) CTS test atlas	69
Bibliography	81
INDEX OF DEFINED TERMS	82
Figure 101 – Electrode position according to Frank (see Table 101).....	38
Figure 102 – Polarity of patient leads (see 51.101.1).....	38
Figure 103 – Test of weighting networks and input impedance (see 51.101.2.2 and 51.102.1) ...	38
Figure 104 – Test of common mode rejection (see 51.105.1 and 51.106.4)	39
Figure 105 – Triangular waveforms for test E of Table 114 (see 51.107.1.1.1).....	40
Figure 106 – Input impulse signal (dashed trace) and cardiograph response (continuous trace) (see 51.107.1.1.2)	40
Figure 107 – Circuit for test of linearity (see 51.107.2)	41
Figure 108 – Result of linearity test (see 51.107.2)	41
Figure 109 – Test of rectangular coordinates (see 51.108.4.1).....	42
Table 101 – ELECTRODES and NEUTRAL ELECTRODES, their position, identification and colour code	13
Table 102 – Offset voltage for ST and T amplitude reference values if the signals are fed through a first order high pass FILTER with a TIME CONSTANT of 3,2 s	16
Table 103 – Acceptable mean differences and standard deviations for global intervals and Q-, R-, S-durations on calibration and analytical ECGs	17
Table 104 – Acceptable mean differences and standard deviations for global durations and intervals for biological ECGs	18
Table 105 – Disclosed changes of measurements caused by NOISE on ECGs according to Table HH.3.....	18
Table 106 – Tabulation of test results	19
Table 107 – Format for disclosure of accuracy measures for diagnostic interpretative statements	22
Table 108 – Format for disclosure of accuracy measures for rhythm interpretative statements...	23
Table 109 – Connection of ELECTRODES for a particular LEAD.....	24
Table 110 – LEADS and their identification (nomenclature and definition).....	24
Table 111 – LEAD networks for Goldberger and Wilson LEADS.....	26
Table 112 – LEAD network for Frank LEADS	27
Table 113 – Test of input impedance – Positions of LEAD SELECTOR, connection of LEAD ELECTRODES and peak to valley deflection in mm with S1 open.....	28
Table 114 – Frequency response	32
Table HH.1 – Calibration and analytical ECGs	66
Table HH.2 – Data set for testing of measurement and wave recognition accuracy of biological data – 100 ECGs of the CSE-study.	67
Table HH.3 – Data set for testing NOISE stability.....	67

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-51: Particular requirements for safety, including essential performance, of recording and analysing single channel and multichannel electrocardiographs

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical specifications, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-51 has been prepared by subcommittee 62D: Electro-medical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this standard is based on the following documents:

FDIS	Report on voting
62D/469/FDIS	62D/473/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

In this Particular Standard, the following print types are used:

- requirements, compliance with which can be tested, and definitions: roman type;
- notes, explanations, advice, introductions, general statements, exceptions and references: small roman type;
- *test specifications: italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD OR THIS PARTICULAR STANDARD: SMALL CAPITALS.

The committee has decided that the contents of this publication will remain unchanged until 2007. At this date, the publication will be

- reconfirmed;
- withdrawn;
- replaced by a revised edition, or
- amended.

A bilingual version of this standard may be issued at a later date.

Withdrawn