

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Magnetic resonance equipment for medical imaging –
Part 1: Determination of essential image quality parameters**

**Appareils à résonance magnétique pour imagerie médicale –
Partie 1: Détermination des principaux paramètres de qualité d'image**





THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2018 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 21 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Catalogue IEC - webstore.iec.ch/catalogue

Application autonome pour consulter tous les renseignements bibliographiques sur les Normes internationales, Spécifications techniques, Rapports techniques et autres documents de l'IEC. Disponible pour PC, Mac OS, tablettes Android et iPad.

Recherche de publications IEC - webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - www.electropedia.org

Le premier dictionnaire en ligne de termes électroniques et électriques. Il contient 21 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Magnetic resonance equipment for medical imaging –
Part 1: Determination of essential image quality parameters**

**Appareils à résonance magnétique pour imagerie médicale –
Partie 1: Détermination des principaux paramètres de qualité d'image**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-6292-4

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	6
INTRODUCTION.....	8
1 Scope.....	10
2 Normative references	11
3 Terms, definitions, symbols and abbreviated terms.....	11
3.1 Terms and definitions.....	11
3.2 Symbols and abbreviated terms	15
4 * Procedures for the determination of essential image parameters.....	16
4.1 General requirements for all procedures	16
4.1.1 Requirements for the system	16
4.1.2 Requirements for the TEST DEVICE.....	16
4.1.3 Scan parameters	16
4.1.4 Reporting of results	17
4.2 * SIGNAL TO NOISE RATIO	20
4.2.1 Objectives and rationale	20
4.2.2 Requirements for the TEST DEVICE.....	20
4.2.3 Scan parameters	20
4.2.4 Measurement procedure	20
4.2.5 Data analysis and tolerances	21
4.2.6 Reporting of results	21
4.3 * UNIFORMITY.....	22
4.3.1 Objectives and rationale	22
4.3.2 Requirements for the TEST DEVICE.....	22
4.3.3 Scan parameters	22
4.3.4 Measurement procedure	23
4.3.5 Data analysis and tolerances	23
4.3.6 Reporting of results	23
4.4 SLICE THICKNESS in 2-D scanning.....	24
4.4.1 Objectives and rationale	24
4.4.2 Requirements for the TEST DEVICE.....	24
4.4.3 Scan parameters	25
4.4.4 Measurement procedure	26
4.4.5 Data analysis and tolerances	26
4.4.6 Reporting of results	27
4.4.7 Reporting of acceptance results.....	27
4.5 * Two-dimensional GEOMETRIC DISTORTION.....	27
4.5.1 Objectives and rationale	27
4.5.2 * Requirements for the TEST DEVICE	28
4.5.3 Scan parameters	30
4.5.4 * Measurement procedure.....	30
4.5.5 * Data analysis and tolerances	31
4.5.6 Reporting of results	32
4.6 * SPATIAL RESOLUTION.....	32
4.6.1 Objectives and rationale	32
4.6.2 Requirements for the TEST DEVICE.....	32
4.6.3 Scan parameters	33
4.6.4 Measurement procedure	34

4.6.5	Data analysis and tolerances	35
4.6.6	Reporting of results	35
4.6.7	Reporting of acceptance results.....	36
4.7	* GHOSTING ARTEFACTS	36
4.7.1	Objectives and rationale	36
4.7.2	* Requirements for the TEST DEVICE	36
4.7.3	Scan parameters	36
4.7.4	Measurement procedure	37
4.7.5	Data analysis and tolerances	37
4.7.6	Reporting of results	38
5	* CONSTANCY TEST	39
5.1	Objectives and rationale	39
5.2	Requirements for the TEST DEVICE	39
5.3	Scan characteristics.....	39
5.4	Measurement procedure	39
5.5	Data analysis, reporting of results and tolerances	40
Annex A	(normative) Alternative methods	41
A.1	Pertaining to 4.2 SIGNAL TO NOISE RATIO.....	41
A.1.1	General	41
A.1.2	Alternative method: SNR measurements using alternative noise determination.....	41
A.1.3	Alternative method: SNR "single image"	42
A.2	Pertaining to 4.3 UNIFORMITY.....	43
A.2.1	General	43
A.2.2	Alternative method "grey-scale map"	43
A.2.3	Alternative method "ACR method"	45
A.3	Pertaining to 4.4 SLICE THICKNESS in 2-D scanning	45
A.3.1	General	45
A.3.2	Alternative method: SLICE THICKNESS in 2-D scanning: wedge method	46
A.4	Pertaining to 4.5 Two-dimensional GEOMETRIC DISTORTION.....	48
A.4.1	General	48
A.4.2	Alternative method: GEOMETRIC DISTORTION measurements using elliptical boundary TEST DEVICES	48
A.4.3	Alternative method: 3D GEOMETRIC DISTORTION component measurement method	49
A.5	Pertaining to 4.6 SPATIAL RESOLUTION	56
A.5.1	General	56
A.5.2	Alternative method: determination of the full MODULATION TRANSFER FUNCTION	57
A.6	Pertaining to 5 CONSTANCY TESTS.....	58
A.6.1	Alternative CONSTANCY TEST methods.....	58
A.6.2	Pitfalls	61
Annex B	(informative) Rationale.....	62
B.1	Pertaining to 4 * Procedures for the determination of essential image parameters	62
B.2	Pertaining to 4.2 SIGNAL TO NOISE RATIO	62
B.2.1	Rationale	62
B.2.2	References	75
B.3	Pertaining to 4.3 UNIFORMITY.....	75
B.3.1	Rationale	75

B.3.2	AAD method	75
B.3.3	Standing waves	75
B.4	Pertaining to 4.5 Two-dimensional GEOMETRIC DISTORTION.....	76
B.4.1	Rationale	76
B.4.2	Pitfalls	77
B.5	Pertaining to 4.6 SPATIAL RESOLUTION.....	81
B.5.1	Rationale	81
B.5.2	Pitfalls	82
B.6	Pertaining to 4.7 GHOSTING ARTEFACTS	83
B.6.1	Rationale	83
B.6.2	Pitfalls	83
B.6.3	References	84
B.7	Pertaining to 5 CONSTANCY TEST – Rationale	84
	Index of defined terms	86
	Bibliography.....	87
	Figure 1 – Signal intensity profile in the inclined slab method	25
	Figure 2 – Correcting for rotation of TEST DEVICE	27
	Figure 3 – Example of a boundary wall TEST DEVICE for a spherical specification volume with two lines passing through the centre.....	29
	Figure 4 – Example of a fiducial marker TEST DEVICE for a spherical specification volume.....	30
	Figure 5 – Distances to be determined.....	31
	Figure 6 – Periodic pattern.....	33
	Figure 7 – Image of periodic pattern and position of ROI for coronal scans	34
	Figure 8 – Image of periodic pattern and position of ROI for transverse and sagittal scans.....	35
	Figure 9 – Example image of the TEST DEVICE and region of interest (ROI) for signal, ghost, and noise measurements.....	38
	Figure A.1 – Wedge TEST DEVICE.....	46
	Figure A.2 – Measurement of SLICE PROFILE and SLICE THICKNESS using wedge TEST DEVICE.....	47
	Figure A.3 – Determination of radius length of an ellipse with semi axis length a and b forming an angle α with respect to the X axis	49
	Figure A.4 – Possible TEST DEVICE configurations for measuring GEOMETRIC DISTORTION.....	51
	Figure A.5 – Two elements with an apparent spacing of $A_i(x,y)$ but a true spacing of $T_i(x,y)$	53
	Figure A.6 – A schematic of a spatial mapping GEOMETRIC DISTORTION plot	55
	Figure A.7 – Scatter plot of GEOMETRIC DISTORTION error	56
	Figure B.1 – Relaxation times T_1 and T_2 in dependency on the concentration of $\text{CuSO}_4 \times 5 \text{H}_2\text{O}$	64
	Figure B.2 – Centring error	78
	Table 1 – Common parameters	18
	Table 2 – Acquisition parameters	19
	Table 3 – Reporting of results for SNR.....	22
	Table 4 – Reporting of results for UNIFORMITY.....	23

Table 5 – Reporting of results for SLICE THICKNESS	27
Table 6 – Reporting of results for GEOMETRIC DISTORTION	32
Table 7 – Phantom, plane and gradient orientation for resolution assessment.....	34
Table 8 – Reporting of results for SPATIAL RESOLUTION.....	36
Table 9 – Reporting of results for GHOSTING ARTEFACTS.....	38
Table 10 – Required CONSTANCY TESTS – Parameter settings	40
Table A.1 – Reporting of results for UNIFORMITY "grey-scale map"	44
Table A.2 – Recommended fiducial volumes	52
Table A.3 – Example of error table.....	56
Table A.4 – Reporting of results for SPATIAL RESOLUTION (MTF method).....	58
Table A.5 – Reporting of results for centre frequency.....	59
Table A.6 – Reporting of results for RF calibration	60
Table A.7 – Reporting of results for geometric accuracy	61
Table B.1 – TEST DEVICE conductivity and dielectric properties	65
Table B.2 – Bandwidth-related quantities as provided by different vendors	67
Table B.3 – Relaxation fit parameters for Gd(TMHD) at concentrations ≤ 4 parts per thousand by weight	68
Table B.4 – Noise correction factors by number of complex channels	69

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MAGNETIC RESONANCE EQUIPMENT FOR MEDICAL IMAGING –**Part 1: Determination of essential image quality parameters**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 62464-1 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This second edition cancels and replaces the first edition published in 2007. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) the tests have been revised to comply with the technical progress;
- b) the range of B_0 was increased from 4 T to 8 T.

The text of this International Standard is based on the following documents:

CDV	Report on voting
62B/1068/CDV	62B/1078/RVC

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements, compliance with which can be tested, and definitions: roman type;
- explanations, advice, notes, general statements and exceptions: smaller roman type;
- *test specifications: italic type;*
- TERMS DEFINED IN CLAUSE 3 OR IN OTHER INTERNATIONAL STANDARDS: SMALL CAPITALS.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex B.

A list of all the parts in the IEC 62464 series, published under the general title *Magnetic resonance equipment for medical imaging*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This part of IEC 62464 is written at a moment in which the magnetic resonance (MR) equipment is already present in the market place for more than 30 years. It is estimated that more than 30 000 scanners are operational and more than 0,5 billion PATIENTS have been scanned. A number of standards on quality assurance and quality control have been developed by National Committees to address the need for quantitative assessment of system performance and system maintenance. It is therefore felt to be necessary to introduce this document in addition to the existing standards on MR safety, because the IEC standards have a true international character and this document combines current best practices together and provides guidance on how to address the various questions on quality control and quality assurance testing of MAGNETIC RESONANCE EQUIPMENT together. Having a standardised set of test methods minimises the amount of work for the MR MANUFACTURERS to demonstrate the performance characteristics of the MR scanners for many different countries and, in addition, these countries do not have to formulate their own requirements for the performance testing.

Since MR scanners have been used for some time, this document is an attempt to consolidate the current way of working for the quality control of the performance characteristics concerning essential image quality parameters, and does not introduce major new development efforts for the established MR EQUIPMENT to fulfil the requirements of this document. This objective is achieved by introducing preferred methods in the main text, while allowing acceptable alternative test methods, described in Annex A. A number of the ACCEPTANCE TEST methods described in this document are already described earlier, mainly as NEMA technical MR standards, and new methods have been developed since then. For this document, it is attempted to select the best method as preferred method, whereas for a number of specific tests, good alternatives are available and are therefore also acceptable.

Also, for the quality assurance tests, the CONSTANCY TESTS, each MANUFACTURER has developed its own TEST DEVICE and related test procedures and data analysis in the past years. For the CONSTANCY TESTS, it was therefore decided not to describe detailed test methods but only prescribe the parameters to be measured and essential conditions for these measurements in the main text. This provides the necessary latitude to account for the many unique MR designs (extremity scanners, whole body scanners, cylindrical versus open scanners, various field strengths, TEST DEVICE design, and data analysis) and examples for possible CONSTANCY TESTS for the required parameters in Annex A.

This document draws on the practical experiences gained in the implementation of IEC 62464-1:2007 and benefits from the continued improvements found in the associated updated NEMA MS series of standards. The utility in implementing the various tests found herein was improved by clarifying the relationship between the tests, the parameters used, the analysis of results, the expected calibration state of the scanner and the reporting of results. Two tests, with no known sensitivity to, for example, field strength considerations (SPATIAL RESOLUTION, SLICE THICKNESS in 2-D scanning), now have acceptance criteria. The Annex A GEOMETRIC DISTORTION test suite now includes 3-D test methods.

An original goal of IEC 62464-1:2007 was linkage of the SNR, SLICE THICKNESS and resolution tests in order to characterize the system in a consistent configuration. However, increasing the range of B_0 covered in this document from 4 T to 8 T (consistent with the recent changes in IEC 60601-2-33) required additional flexibility in the TEST DEVICE filler composition in order to eliminate confounding wavelength ARTEFACTS. Therefore, the various test clauses are decoupled in this document. This permits the flexibility to perform each test in an optimal configuration and does not require a retest of other parameters. For example, it is not necessary to repeat a resolution test for a RF COIL, which is not a function of the RF COIL, when the objective is to measure only SNR.

It was not possible to establish a full set of TEST DEVICE and scan parameter requirements appropriate for all MRI systems at the full range of B_0 permitted in this document. Instead, this document was modified to state that testing shall be performed in an MRI system that has been properly calibrated for routine clinical scanning. Calibration is specific to the make and model of MRI scanner and no requirements are listed in this document. The flexibility in the

definition of specification areas and volumes was improved to support the increasing specialization of receive coils. The standard encourages reuse of phantoms for multiple applications where possible, as long as the phantom provides signal in the specification area and/or volume as required, unless instructed otherwise.

This document has also been modified regarding the use of reconstruction and image filters. The intent of IEC 62464-1:2007 was to disable all user controlled filters, and record the condition of all other filters, in order to characterize the system in the most basic possible configuration. However, systems continue to evolve and the presence and configuration of some filters are not known to the end user (e.g. image reconstruction), whereas other filters might be known to the end user, beyond their control, and always applied (e.g. GEOMETRIC DISTORTION correction). This document formally introduces two mechanisms for addressing this situation: 1) the concept of "clinically relevant" to provide guidance on filter settings and 2) an emphasis on the accurate recording of all parameters used in the acquisition and reconstruction of the images sufficient to enable a faithful reproduction of results on another unit of the same make, model and software revision. The intent of "clinically relevant" is to enable a known and properly identified protocol from a given software revision to be used as the basis for the tests. Factory set defaults are assumed to be applied, and filters considered not essential can be turned on or off as clinically appropriate. All adjustments made from default setting should be recorded in the reporting of results. By carefully recording the base system configuration and any additional acquisition parameter adjustments, all known and unknown filter settings are reproducible and all results should be repeatable. Note that "clinically relevant" also enables the user of this document to appropriately select parameters (e.g. acquisition bandwidth) that may vary with B_0 or other system attributes.

MAGNETIC RESONANCE EQUIPMENT FOR MEDICAL IMAGING –

Part 1: Determination of essential image quality parameters

1 Scope

This part of IEC 62464 specifies measurement procedures for the determination of many essential image quality parameters for MR EQUIPMENT. Measurement procedures as addressed in this document are suitable for

- quality assessment in the ACCEPTANCE TEST, and
- quality assurance in the CONSTANCY TEST.

Required levels of performance for ACCEPTANCE TESTS are not provided for all tests.

This document does not address

- image quality assessment of MR EQUIPMENT with a static magnetic field intensity greater than 8 Tesla, if not otherwise stated,
- image quality affected by MR-compatibility issues,
- special diagnostic procedures such as flow imaging, perfusion, diffusion, radiotherapy and image-guided therapy applications, and
- TYPE TESTS.

The scope of this document is also limited to measuring image quality characteristics in images acquired on TEST DEVICES, not in PATIENT images.

The measurement procedures specified in this document are directed to

- MANUFACTURERS, who can demonstrate compliance by performing ACCEPTANCE and CONSTANCY TESTS as described by this document,
- test houses, who can confirm performance of MR EQUIPMENT using methods described in this document,
- regulatory authorities, who can reference this document, and
- RESPONSIBLE ORGANISATIONS who want to perform ACCEPTANCE and CONSTANCY TESTS using methods described in this document.

The essential image quality parameters and measurement methodologies defined in this document are

- SIGNAL TO NOISE RATIO,
- UNIFORMITY,
- SLICE THICKNESS in 2-D scanning,
- 2-D GEOMETRIC DISTORTION,
- SPATIAL RESOLUTION, and
- GHOSTING ARTEFACTS.

Each of these procedures can be performed standalone or in combination with any of the other procedures.

This document describes the preferred measurement procedures. It also describes alternative normative methods in Annex A. The preferred test methods may be substituted with these

alternative normative methods. If necessary, other methods not described in this document can be used, provided those other test methods are documented and validated against the methods described in the document: it means an analysis is done by comparison to the original method that demonstrates a similar, or better, level of sensitivity to the same parameter of interest and a similar, or better, level of robustness against unrelated parameters. All methods will produce quantitative results.

The rationale to the preferred and alternate methods, and their pitfalls, are described in Annex B.

This document also presents requirements for CONSTANCY TESTS suitable for MR EQUIPMENT quality assurance programs concerning essential image quality parameters. There are no preferred CONSTANCY TEST methods, to provide flexibility in using existing automated procedures where available, but suggested examples of test methods can be found in Annex A. This document places an emphasis on consistently repeatable, automated measuring tools that facilitate trend analysis and the frequent quick testing of a small set of important parameters that are sensitive to the overall operating characteristics of the MR EQUIPMENT.

NOTE None of the methods found in this document have been extensively tested at a static magnetic field intensity above 3 T. Initial tests indicate the methods function correctly when appropriate TEST DEVICE fillers are used.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2013

IEC 60601-2-33:2010, *Medical electrical equipment – Part 2-33: Particular requirements for the basic safety and essential performance of magnetic resonance equipment for medical diagnosis*

IEC TR 60788, *Medical electrical equipment – Glossary of defined terms*

3 Terms, definitions, symbols and abbreviated terms

3.1 Terms and definitions

For the purposes of this document, the terms and definitions in IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, IEC 60601-2-33, IEC TR 60788 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE 1 See also the IEC Glossary available at <http://std.iec.ch/glossary>.

NOTE 2 An index of defined terms is found beginning on page 86.

3.1.1

ACCEPTANCE TEST

test carried out after new equipment has been installed, or major modifications have been made to existing equipment, in order to verify compliance with contractual specifications

[SOURCE: IEC 61223-3-5:2004, 3.1]

3.1.2

ARTEFACT

apparent structure visible in the image that does not represent a structure at the corresponding position within the object and that cannot be explained by noise

3.1.3

BANDWIDTH PER PIXEL

BPP

reciprocal of the duration of the sampling window

Note 1 to entry: Instead of BANDWIDTH PER PIXEL, alternate values may be stated on the control console in a MANUFACTURER-specific format. For a rationale on how to convert the MANUFACTURER-specific information into the required BANDWIDTH PER PIXEL, see B.2.1.5.

Note 2 to entry: This definition is only applicable when standard, constant-gradient readout is used.

Note 3 to entry: This note applies to the French text only.

3.1.4

CONSTANCY TEST

each of a series of tests carried out to ensure that the functional performance of equipment meets established criteria; or to enable the early recognition of changes in the properties of components of the equipment

[SOURCE: IEC 61223-3-5:2004, 3.2]

3.1.5

EDGE SPREAD FUNCTION

ESF

discrete profile of the image data, taken across a sharp edge

Note 1 to entry: This note applies to the French text only.

3.1.6

FIELD OF VIEW

FOV

size of the imaging area requested by the OPERATOR

Note 1 to entry: To be specified with one or two linear measures (in mm), if imaging area is square or rectangular respectively.

3.1.7

FULL WIDTH AT HALF-MAXIMUM

FWHM

interval parallel to the abscissa between the points on a curve with the value of one-half of the maximum of the curve

[SOURCE: IEC 61223-2-6:2004, 3.9]

3.1.8

GEOMETRIC DISTORTION

spatial position deviation of the imaged structure from expected position of real object structure

3.1.9**GHOSTING ARTEFACT**

ARTEFACT showing a replica or part of a replica of an existing structure in a wrong position

3.1.10**IMAGE NOISE**

amplitude of the random fluctuations from the expected signal values in an image

3.1.11**ISOCENTRE**

null point of the spatially encoding gradients

Note 1 to entry: Typically, this also corresponds to the region of highest magnet homogeneity.

3.1.12**LINE SPREAD FUNCTION****LSF**

complex derivative of the EDGE SPREAD FUNCTION

Note 1 to entry: This note applies to the French text only.

3.1.13**MAGNETIC RESONANCE EQUIPMENT****MR EQUIPMENT**

MEDICAL ELECTRICAL EQUIPMENT which is intended for in vivo magnetic resonance examination of a PATIENT comprising all parts in hardware and software from the SUPPLY MAINS to the display monitor

Note 1 to entry: The MR EQUIPMENT is a PROGRAMMABLE ELECTRICAL MEDICAL SYSTEM (PEMS).

[SOURCE: IEC 60601-2-33:2010, 201.3.218]

3.1.14**MANUFACTURER**

natural or legal person with responsibility for the design, manufacture, packaging, or labelling of ME EQUIPMENT, assembling an ME SYSTEM, or adapting ME EQUIPMENT or an ME SYSTEM, regardless of whether these operations are performed by that person or on that person's behalf by a third party

[SOURCE: IEC 60601-1:2005, 3.55, modified – The notes have been deleted.].

3.1.15**MODULATION TRANSFER FUNCTION****MTF**

normalised magnitude of the Fourier transform of the LINE SPREAD FUNCTION

3.1.16**OPERATOR**

person handling equipment

[SOURCE: IEC 60601-1:2005, 3.73]

3.1.17**PATIENT**

living being (person or animal) undergoing a medical, surgical or dental procedure

Note 1 to entry: A PATIENT can be an OPERATOR.

[SOURCE: IEC 60601-1:2005, 3.76, modified – The note has been added.]

3.1.18**PHASE OVERSAMPLING**

acquisition technique increasing by the same factor both the phase-encoding FIELD OF VIEW and the number of phase encodings, to avoid aliasing ARTEFACTS in the phase encoding direction

Note 1 to entry: The additional acquired part of the phase-encoding FIELD OF VIEW is cropped away.

Note 2 to entry: The technique has various commercial names, such as "oversampling" (Esaote), "no phase-wrap" (GE), "anti-wrap" (Hitachi), "fold-over suppression" (Philips), "PHASE OVERSAMPLING" (Siemens), "phase-wrap suppression" (Toshiba). This information is given for the convenience of users of this document and does not constitute an endorsement by IEC of the companies named.

3.1.19**RADIO FREQUENCY COIL****RF COIL**

coil or probe used to transmit and/or receive radio frequency electromagnetic fields

3.1.20**RECONSTRUCTION PARAMETERS**

ensemble of all parameters that uniquely define the transformation of the digitized signals into an image

3.1.21**REFERENCE POSITION**

predefined point within the SPECIFICATION VOLUME

Note 1 to entry: Normally, the REFERENCE POSITION lies on the symmetry axis of the RF COIL and corresponds to the centre of gravity of the SPECIFICATION VOLUME.

3.1.22**REGION OF INTEREST****ROI**

localised part of an image, which is of particular interest at a given time

Note 1 to entry: This note applies to the French text only.

[SOURCE: IEC 61223-3-5:2004, 3.15]

3.1.23**RESPONSIBLE ORGANISATION**

entity accountable for the use and maintenance of an ME EQUIPMENT or an ME SYSTEM

Note 1 to entry: The accountable entity can be, for example, a hospital, an individual clinician or a layperson. In home use applications, the PATIENT, OPERATOR and RESPONSIBLE ORGANIZATION can be one and the same person.

Note 2 to entry: Education and training is included in "use".

[SOURCE: IEC 60601-1:2005, 3.101]

3.1.24**RF COIL SPECIFICATION AREA**

intersection of the RF COIL SPECIFICATION VOLUME and the image plane

3.1.25**RF COIL SPECIFICATION VOLUME**

imaging volume, defined relative to a specific RF COIL, within which a MANUFACTURER characterizes image performance

Note 1 to entry: Each RF COIL has its own appropriately shaped specification volume.

3.1.26**SIGNAL TO NOISE RATIO****SNR**

ratio of the signal level to the noise level

3.1.27**SLICE PROFILE**

plot of the signal intensity in a direction perpendicular to the scanned slice

3.1.28**SLICE THICKNESS**

FULL WIDTH AT HALF MAXIMUM of the SLICE PROFILE

3.1.29**SPATIAL RESOLUTION**

one half of the inverse of the highest spatial frequency where the magnitude of the MODULATION TRANSFER FUNCTION (MTF) exceeds a required threshold

3.1.30**SYSTEM SPECIFICATION VOLUME**

imaging volume within which a MANUFACTURER characterises image performance

Note 1 to entry: Images or portions of images outside this volume will not necessarily meet performance specifications, but can still be useful for diagnostic purposes.

Note 2 to entry: The SYSTEM SPECIFICATION VOLUME may be different per image quality parameter.

3.1.31**TEST DEVICE**

object for imaging test purposes

3.1.32**TYPE TEST**

test on a representative sample of the equipment with the objective of determining if the equipment, as designed and manufactured, can meet the requirements of IEC 60601-1

[SOURCE: IEC 60601-1:2005, 3.135]

3.1.33**UNIFORMITY**

spatial constancy of intensity in the image of a homogeneous TEST DEVICE excluding the effect of noise or finite data collection (truncation ARTEFACT)

3.2 Symbols and abbreviated terms

B_0	static magnetic field
B_1	radio frequency magnetic field
FT	Fourier transform
T_1	longitudinal relaxation time
T_2	transversal relaxation time
T_E	echo time
T_R	repetition time