

INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 2-091: Particular requirements for cabinet X-ray systems**

**Règles de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire –
Partie 2-091: Exigences particulières pour les équipements à rayons X montés en armoire**





THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2019 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

Glossaire IEC - std.iec.ch/glossary

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.



INTERNATIONAL STANDARD

NORME INTERNATIONALE

**Safety requirements for electrical equipment for measurement, control, and laboratory use –
Part 2-091: Particular requirements for cabinet X-ray systems**

**Règles de sécurité pour appareils électriques de mesurage, de régulation et de laboratoire –
Partie 2-091: Exigences particulières pour les équipements à rayons X montés en armoire**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 19.080; 71.040.10

ISBN 978-2-8322-6518-5

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD	3
INTRODUCTION	6
1 Scope and object	7
2 Normative references	8
3 Terms and definitions	8
4 Tests	9
5 Marking and documentation	10
6 Protection against electric shock	13
7 Protection against mechanical HAZARDS	15
8 Resistance to mechanical stresses	16
9 Protection against the spread of fire	16
10 Equipment temperature limits and resistance to heat	16
11 Protection against HAZARDS from fluids and solid foreign objects	16
12 Protection against radiation, including laser sources, and against sonic and ultrasonic pressure	16
13 Protection against liberated gases and substances, explosion and implosion	19
14 Components and subassemblies	19
15 Protection by INTERLOCKS	20
16 HAZARDS resulting from application	22
17 RISK assessment	22
Annexes	23
Annex L (informative) Index of defined terms	24
Annex AA (normative) Standard phantom for PARTIALLY PROTECTED EQUIPMENT	25
Annex BB (informative) Dose limit recommendation for occupational radiation	26
Bibliography	27
Figure AA.1 – Phantom for measurement of PARTIALLY PROTECTED EQUIPMENT	25
Table 101 – Durations of high voltage strength test	14
Table BB.1 – Dose limits	26

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT
FOR MEASUREMENT, CONTROL, AND LABORATORY USE –****Part 2-091: Particular requirements for cabinet X-ray systems**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61010-2-091 has been prepared by IEC technical committee 66: Safety of measuring, control and laboratory equipment.

This second edition cancels and replaces the first edition published in 2012. It constitutes a technical revision.

This edition includes the following significant changes from the first edition, as well as numerous other changes:

- The scope of the document has been clarified and limited to equipment up to 500 kV.
- Additional marking requirements for X-ray generating assemblies have been added. (5.1)
- Requirements for high-voltage cables used in the X-ray assembly have been added. (6.5)
- Insulation requirements have been added. (6.7)
- Temperature requirements for beam-limiting devices have been added. (10.3)

- Clarification has been provided on PROTECTED EQUIPMENT and PARTIALLY PROTECTED EQUIPMENT, and test methods. (12)
- Requirements for INTERLOCKS have been modified, taking into account functional safety standards. (15)
- Requirements for reasonably foreseeable misuse have been clarified. (16)
- Risk assessment has been made mandatory for specific aspects. (17)

The text of this International Standard is based on the following documents:

FDIS	Report on voting
66/684/FDIS	66/686A/RVD

Full information on the voting for the approval of this International Standard can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

This document is intended to be used in conjunction with IEC 61010-1. It was established on the basis of the third edition (2010) of IEC 61010-1, including its Amendment 1 (2016), hereinafter referred to as Part 1.

This Part 2-091 supplements or modifies the corresponding clauses in IEC 61010-1 so as to convert that publication into the IEC standard: *Particular requirements for cabinet X-ray systems*.

Clauses of Part 1 that are fully applicable are indicated by the statement "This clause of Part 1 is applicable." Where this Part 2-091 identifies a particular subclause and states "addition", "modification", "replacement", or "deletion", the text of that particular subclause Part 1 is adapted accordingly. Where a particular subclause of Part 1 is not mentioned in this Part 2-091, that subclause applies as far as is reasonable.

In this standard:

- a) the following print types are used:
 - requirements: in roman type;
 - NOTES: in small roman type;
 - conformity and tests: *in italic type*;
 - terms used throughout this standard which have been defined in Clause 3: SMALL ROMAN CAPITALS.
- b) subclauses, figures, and tables which are additional to those in Part 1 are numbered starting from 101. Additional annexes are lettered starting from AA and additional list items are lettered from aa).

A list of all parts of the IEC 61010 series, published under the general title *Safety requirements for electrical equipment for measurement, control, and laboratory use*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

IEC 61010-1 specifies the safety requirements that are generally applicable to all equipment within its scope. For certain types of equipment, the requirements of IEC 61010-1 and its amendments will be supplemented or modified by the special requirements of one, or more than one, particular Part 2s of the standard, which are to be read in conjunction with the Part 1 requirements.

This document has been prepared, based on IEC 61010-1:2010 including its Amendment 1:2016, to specify additional safety requirements for cabinet X-ray systems. It provides additional guidance for construction and assessment of extra high voltage circuits, mechanical HAZARDS and ionizing radiation HAZARDS which can be present in this type of equipment.

This document has been written to provide protection against both radiation HAZARDS from the direct X-ray beam and any scattered X-radiation caused by reflections of the X-ray beam on any part of the equipment or on the sample subjected to X-rays.

The minimum safety requirements specified in this document are considered to provide for a practical degree of safety in the operation of cabinet X-ray systems.

SAFETY REQUIREMENTS FOR ELECTRICAL EQUIPMENT FOR MEASUREMENT, CONTROL, AND LABORATORY USE –

Part 2-091: Particular requirements for cabinet X-ray systems

1 Scope and object

This clause of Part 1 is applicable, except as follows:

1.1 Scope

1.1.1 Equipment included in scope

Deletion:

Delete the first paragraph.

Replacement:

Replace the second paragraph (above items a) to c)) with the following new text:

This part of IEC 61010 specifies particular safety requirements for cabinet X-ray systems, which fall under any of categories a), b) or c) below.

Addition:

Add the two following new paragraphs at the end of the subclause:

Equipment covered by this document can be both PROTECTED EQUIPMENT or PARTIALLY PROTECTED EQUIPMENT, with X-ray generator voltage up to 500 kV.

A cabinet X-ray system is a system that contains an X-ray tube installed in a cabinet, which, independently of existing architectural structures except the floor on which it may be placed, is intended to contain at least that portion of a material being irradiated, provide radiation attenuation and prevent operator access to the radiation beam, during generation of X-radiation.

These cabinet X-ray systems are used in industrial, commercial, and public environments, for example, to inspect materials, to analyse materials, and to screen baggage.

1.1.2 Equipment excluded from scope

Addition:

Add the following new items to the list:

- aa) Equipment intended to apply X-radiation to humans or animals;
- bb) Equipment incorporating an X-ray tube but not incorporating complete shielding against X-radiation HAZARDS, such as:
 - equipment intended to be used within a shielded room which excludes personnel during operation;
 - equipment intended to be used with separate portable or temporary shielding;
 - equipment intended to produce an emerging beam of X-radiation.

1.2 Object

1.2.1 Aspects included in scope

Addition:

Add the following new text to the end of the first paragraph:

This part of IEC 61010 specifies requirements for the design and methods of construction of cabinet X-ray systems to provide adequate protection for OPERATORS, bystanders, trained service personnel and the surrounding area against unintentionally-emitted X-radiation and from mechanical HAZARDS related to their conveyors.

2 Normative references

This clause of Part 1 is applicable, except as follows:

Addition:

Add the following references to the list:

IEC 62061, *Safety of machinery – Functional safety of safety-related electrical, electronic and programmable electronic control systems*

ISO 13849-1, *Safety of machinery – Safety-related parts of control systems – Part 1: General principles for design*

3 Terms and definitions

This clause of Part 1 is applicable, except as follows:

3.2 Parts and accessories

Addition:

Add the following new terms and definitions:

3.2.101

ACCESS PANEL

PROTECTIVE BARRIER or panel which is designed to be removed or opened through the use of a TOOL for maintenance or service purposes to permit access to the interior of the cabinet

3.2.102

APERTURE

opening in the outside surface of the cabinet, other than a PORT, which remains open during generation of X-radiation

3.2.103

DOOR

PROTECTIVE BARRIER which is designed to be movable or opened for routine operation purposes, does not generally require TOOLS to open and permits access to the interior of the cabinet

Note 1 to entry: Inflexible hardware rigidly affixed to the DOOR is considered part of the DOOR.

Note 2 to entry: Access openings intended for the OPERATOR, for removal or re-alignment of samples, are considered as a DOOR.

3.2.104**EXTERNAL SURFACE**

outside surface of the cabinet X-ray system, including DOORS, ACCESS PANELS, latches, control knobs and other permanently mounted hardware, the virtual surface across any APERTURE or PORT, and the bottom of the cabinet

3.2.105**PORT**

opening in the EXTERNAL SURFACE of the cabinet which is designed to remain open during generation of X-rays, for the purpose of conveying objects into and out of the cabinet or for partial insertion for irradiation of an object with dimensions that do not permit complete insertion into the cabinet

3.2.106**PROTECTED EQUIPMENT**

cabinet X-ray system without any APERTURE or PORT, which would allow access to any area with X-radiation

Note 1 to entry: Access can be prevented through INTERLOCK protected DOORS, PORTS, APERTURES or ACCESS PANELS.

3.2.107**PARTIALLY PROTECTED EQUIPMENT**

cabinet X-ray system with APERTURE or PORT, allowing access to any area with X-radiation

Note 1 to entry: Access to the X-ray beam or scattered X-radiation could be prevented through a combination of ENCLOSURE and flexible radiation absorbers, for example, one or several curtains containing lead or another radiation absorbing material.

3.2.108**BEAM-LIMITING DEVICE**

device to limit the radiation field, such as a collimator, a cone or an APERTURE, intended to restrict the dimensions of X-ray field

3.2.109**INTERLOCK**

arrangement of components or devices operating together, intended to prevent a HAZARD or specific operation whenever safety is compromised by access to the interior of the system, operational irregularity or equipment failure

[SOURCE: IEC 62463:2010, 3.11, modified – The term "safety interlocks" has been replaced with "interlock" and the wording "interrupt the generation of X-radiation" has been replaced with "a HAZARD or specific operation".]

4 Tests

This clause of Part 1 is applicable, except as follows:

4.3.2.4 Covers and removable parts

Addition:

Add the following new paragraph at the end of the subclause:

For the test in 12.2.1.101.3, any flexible radiation absorbers used to close the APERTURE or PORT openings for PARTIALLY PROTECTED EQUIPMENT shall remain in their normal closed position.