

Essais d'évaluation et de routine dans les services d'imagerie médicale –

Evaluation and routine testing in medical imaging departments –

Partie 3-5: Essais d'acceptation - Performance d'imagerie des équipements de tomodensitométrie à rayonnement X

Part 3-5: Acceptance tests – Imaging performance of computed tomography X-ray equipment

CORRIGENDUM 1

Page 56

F.4 Evaluation des données avec le DISPOSITIF D'ESSAI alternatif «mire à barres parallèles»

Remplacer l'équation existante:

$$FTM = \frac{M_o}{M_i} = \frac{\pi}{\sqrt{2}} \times \frac{\sqrt{M_{\text{mire}} - N_{\text{fond}}}}{|CT_{\text{matériau}} - CT_{\text{eau}}|}$$

par la nouvelle équation suivante:

$$FTM = \frac{M_o}{M_i} = \frac{\pi}{\sqrt{2}} \times \frac{\sqrt{M_{\text{mire}}^2 - N_{\text{fond}}^2}}{|CT_{\text{matériau}} - CT_{\text{eau}}|}$$

Page 62

Bibliographie

Au lieu de :

[8] JUDY, P. F. The line spread function and modulation transfer function of a computed tomographic scanner. *J Comput Assist Tomogr* 3, 1976, p.189-195.

lire:

[8] JUDY, P. F. The line spread function and modulation transfer function of a computed tomographic scanner. *Medical Physics*, July, 1976, Vol. 3, Issue 4, p. 233-236

Page 57

F.4 Data evaluation with the alternate TEST DEVICE "bar pattern"

Replace the existing equation:

$$MTF = \frac{M_o}{M_i} = \frac{\pi}{\sqrt{2}} \times \frac{\sqrt{M_{\text{pattern}} - N_{\text{background}}}}{|CT_{\text{material}} - CT_{\text{water}}|}$$

with the following new equation:

$$MTF = \frac{M_o}{M_i} = \frac{\pi}{\sqrt{2}} \times \frac{\sqrt{M_{\text{pattern}}^2 - N_{\text{background}}^2}}{|CT_{\text{material}} - CT_{\text{water}}|}$$

Page 63

Bibliography

Instead of:

[8] JUDY, P. F. The line spread function and modulation transfer function of a computed tomographic scanner. *J Comput Assist Tomogr* 3, 1976, p.189-195.

read:

[8] JUDY, P. F. The line spread function and modulation transfer function of a computed tomographic scanner. *Medical Physics*, July, 1976, Vol. 3, Issue 4, p. 233-236