

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE

**Radiation protection instrumentation – Installed ambient dose equivalent rate meters, warning and monitoring assemblies for neutrons with energies from thermal to 20 MeV**

**Instrumentation pour la radioprotection – Débitmètres d'équivalent de dose ambiant, ensembles d'alarmes et moniteurs à poste fixe pour des énergies de neutrons comprises entre l'énergie thermique et 20 MeV**





## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2020 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 000 terminological entries in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

#### IEC Glossary - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

[webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études,...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 000 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 16 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

#### Glossaire IEC - [std.iec.ch/glossary](http://std.iec.ch/glossary)

67 000 entrées terminologiques électrotechniques, en anglais et en français, extraites des articles Termes et Définitions des publications IEC parues depuis 2002. Plus certaines entrées antérieures extraites des publications des CE 37, 77, 86 et CISPR de l'IEC.

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

---

**Radiation protection instrumentation – Installed ambient dose equivalent rate meters, warning and monitoring assemblies for neutrons with energies from thermal to 20 MeV**

**Instrumentation pour la radioprotection – Débitmètres d'équivalent de dose ambiant, ensembles d'alarmes et moniteurs à poste fixe pour des énergies de neutrons comprises entre l'énergie thermique et 20 MeV**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

---

ICS 13.280; 17.240

ISBN 978-2-8322-7635-8

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

# CONTENTS

FOREWORD ..... 6

1 Scope ..... 8

2 Normative references ..... 9

3 Terms and definitions, abbreviated terms and symbols, quantities and units ..... 10

    3.1 Terms and definitions ..... 10

    3.2 Abbreviated terms and symbols ..... 17

    3.3 Quantities and units ..... 18

4 Design requirements ..... 18

    4.1 Construction and performance ..... 18

        4.1.1 General ..... 18

        4.1.2 Equipment configuration ..... 18

        4.1.3 Equipment reliability ..... 18

    4.2 Indication facilities ..... 19

    4.3 Alarm facilities ..... 19

        4.3.1 General ..... 19

        4.3.2 High level alarms ..... 20

        4.3.3 Low level alarms ..... 20

        4.3.4 Instrument fault alarms ..... 20

    4.4 External facilities ..... 20

    4.5 Effective range of measurement ..... 20

        4.5.1 General ..... 20

        4.5.2 Requirements ..... 21

        4.5.3 Interrelationship between response time and statistical fluctuations ..... 21

        4.5.4 Rated range of an influence quantity ..... 21

        4.5.5 Minimum rated range of influence quantity ..... 21

    4.6 Ingress protection (IP) classification ..... 21

    4.7 Assembly labels and markings ..... 22

    4.8 Functional safety ..... 22

5 General test procedure ..... 22

    5.1 Test requirements ..... 22

    5.2 Tests performed with variation of influence quantities ..... 23

        5.2.1 General ..... 23

        5.2.2 Tests for influence quantities of type F ..... 25

        5.2.3 Tests for influence quantities of type S ..... 25

    5.3 Point of test ..... 25

    5.4 Statistical fluctuations ..... 25

    5.5 Radiation sources ..... 26

    5.6 Work place neutron fields ..... 26

    5.7 Consideration of several detectors or signals in a dose rate meter ..... 26

    5.8 Functionality test ..... 27

        5.8.1 General ..... 27

        5.8.2 Pre-test (preparation) ..... 27

        5.8.3 Post-test ..... 27

6 Radiation detection requirements ..... 28

    6.1 General ..... 28

    6.2 Consideration of the uncertainty of the conventional true value ..... 28

6.3	Linearity of dose rate response .....	28
6.3.1	Requirements .....	28
6.3.2	Tests to be performed.....	28
6.3.3	Type tests.....	28
6.3.4	Routine tests .....	29
6.3.5	Test method using sources .....	29
6.3.6	Interpretation of the results of the test using sources .....	29
6.3.7	Test procedure with variation of the calibration distance .....	29
6.3.8	Interpretation of the results of the test with variation of the calibration distance.....	29
6.3.9	Equivalent electrical test method .....	30
6.3.10	Interpretation of the equivalent electrical test result .....	30
6.4	Variation of the response due to neutron energy .....	30
6.4.1	General .....	30
6.4.2	Requirements .....	30
6.4.3	Method of test.....	31
6.4.4	Interpretation of the results .....	32
6.5	Variation of response with angle of incidence.....	32
6.5.1	General .....	32
6.5.2	Requirements .....	32
6.5.3	Method of test.....	34
6.5.4	Interpretation of the results of the test .....	35
6.6	Response to other external ionizing radiations .....	35
6.6.1	Alpha and beta radiations .....	35
6.6.2	Photon radiation .....	35
6.7	Overload characteristics .....	36
6.7.1	Requirements .....	36
6.7.2	Method of test.....	36
6.8	Statistical fluctuations .....	37
6.8.1	Requirements .....	37
6.8.2	Method of test.....	37
6.8.3	Interpretation of the results.....	37
6.9	Response time .....	37
6.9.1	Requirements .....	37
6.9.2	Method of test.....	38
6.10	Zero drift.....	38
6.10.1	Requirements .....	38
6.10.2	Method of test.....	38
6.11	Alarm requirements.....	38
6.11.1	Requirements .....	38
6.11.2	Method of test.....	39
6.12	Equipment failure alarms .....	39
6.12.1	Requirements .....	39
6.12.2	Method of test.....	39
6.13	False alarm test .....	39
6.13.1	Requirements .....	39
6.13.2	Method of test.....	39
6.14	Alarm response time and stability .....	39
6.14.1	Requirements .....	39

6.14.2	Method of test.....	39
6.15	Warm-up.....	40
6.15.1	Requirements .....	40
6.15.2	Method of test.....	40
6.16	Power supplies .....	41
6.16.1	Requirements .....	41
6.16.2	Method of test.....	41
7	Environmental requirements .....	41
7.1	General.....	41
7.2	Ambient temperature.....	42
7.2.1	Requirements .....	42
7.2.2	Method of test.....	42
7.3	Relative humidity .....	42
7.3.1	Requirements .....	42
7.3.2	Method of test.....	42
7.4	IP classification.....	43
7.4.1	Requirements .....	43
7.4.2	Method of test.....	43
8	Mechanical requirements.....	43
8.1	General.....	43
8.2	Microphonics/impact .....	43
8.2.1	Requirements .....	43
8.2.2	Method of test.....	43
9	Electromagnetic requirements .....	44
9.1	General.....	44
9.2	Emission of electromagnetic radiation .....	45
9.2.1	Requirements .....	45
9.2.2	Method of test.....	45
9.3	Electrostatic discharge.....	45
9.3.1	Requirements .....	45
9.3.2	Method of test.....	45
9.4	Radio frequency disturbance.....	46
9.4.1	Requirements .....	46
9.4.2	Method of test.....	46
9.5	Magnetic fields.....	46
9.5.1	Requirements .....	46
9.5.2	Method of test.....	46
9.6	Alternating current powered equipment requirements.....	47
9.6.1	Voltage and frequency fluctuations .....	47
9.6.2	Immunity from conducted RF .....	47
9.6.3	Surges and ring waves .....	48
10	Documentation .....	48
10.1	Type test report .....	48
10.2	Certificate .....	48
10.3	Operation and maintenance manual.....	49
Annex A (informative) Neutron fluence-to-ambient dose equivalent conversion coefficients .....		50
Bibliography.....		53

Figure 1 – Test for angular response .....	33
Figure A.1 – Neutron fluence-to-ambient dose equivalent conversion coefficients for mono-energetic neutrons [3] .....	51
Table 1 – Reference and standard test conditions .....	23
Table 2 – Radiation characteristics of ambient neutron dose equivalent rate meters .....	24
Table 3 – Electrical and environmental characteristics of ambient dose equivalent rate meters .....	40
Table 4 – Maximum values of deviation due to electromagnetic disturbances.....	44
Table 5 – Emission frequency range .....	45
Table A.1 – Neutron fluence-to-ambient dose equivalent conversion coefficients for mono-energetic neutrons ([1], [3], [4], [7]) .....	50
Table A.2 – Neutron fluence-to-ambient dose equivalent conversion coefficients for the neutron reference radiation sources ([3] and ISO 8529-3) .....	52

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**RADIATION PROTECTION INSTRUMENTATION –  
INSTALLED AMBIENT DOSE EQUIVALENT RATE METERS,  
WARNING AND MONITORING ASSEMBLIES FOR NEUTRONS  
WITH ENERGIES FROM THERMAL TO 20 MeV**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 61322 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation.

This second edition cancels and replaces the first edition published in 1994. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- this document has been updated to take account of the requirements of the relevant IEC standards, IEC 60532:2010 and IEC 61005:2014.

The text of this document is based on the following documents:

FDIS	Report on voting
45B/944/FDIS	45B/952/RVD

Full information on the voting for the approval of this document can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

# RADIATION PROTECTION INSTRUMENTATION – INSTALLED AMBIENT DOSE EQUIVALENT RATE METERS, WARNING AND MONITORING ASSEMBLIES FOR NEUTRONS WITH ENERGIES FROM THERMAL TO 20 MeV

## 1 Scope

This document applies to installed dose equivalent rate meters, warning assemblies and monitors, as defined below. It covers equipment intended to measure neutron radiation in dose equivalent rates in the energy region between thermal and 20 MeV for the purposes of radiation protection.

Assemblies of this type are commonly defined as area radiation monitors. They are normally employed to determine continuously the radiological situation in working areas in which the radiation field may change with time, for example, nuclear power plants, particle accelerators, high-activity laboratories, fuel reprocessing plants, etc., and provide alarms when the radiation field goes outside predetermined limits.

The assemblies considered in this document comprise at least:

- a detector assembly, which may, for example, consist of a detector probe (for thermal neutrons such as BF<sub>3</sub> proportional counter, <sup>3</sup>He proportional counter, <sup>6</sup>LiI(Tl) scintillation detector, etc.) and a moderating and absorbing medium surrounding the detector;
- a processing assembly, which may be fitted into a centralized panel, which, in the case of warning assemblies and monitors, provides signal outputs and contacts capable of activating alarm or other trip circuits;
- alternatively, the case when all the processing electronics are placed within the detection unit (so called "smart blocks") may be considered. In this case the functions of the processing assembly will be composed of only the indication, the providing signal outputs and contacts.

This document specifies general characteristics, general test procedures, radiation characteristics, as well as electrical, mechanical, safety, and environmental characteristics, as well as the identification certificate for the assemblies defined in the scope.

Assemblies designed to perform combined functions comply with the requirements pertaining to each of these functions.

This document is not applicable to criticality monitors covered by IEC 60860, or to assemblies intended to give information about operational parameters of nuclear plants for control purposes. This document is not applicable to the operating characteristics of indicating or recording instruments as such (for instance, indicating meters, recorders, etc.). The characteristics of such instruments are in conformity with the general requirements appropriate to them.

This document does not cover hand-held neutron dose (rate) meters and instruments that are covered in IEC 61005.

No tests are specified in this document for performance requirements in pulsed radiation fields. It is understood that an assembly designed to meet this document may not be suitable for use in these fields.