

TECHNICAL SPECIFICATION

Ultrasonics – Measurements of electroacoustical parameters and acoustic output power of spherically curved transducers using the self-reciprocity method

Withdrawn



THIS PUBLICATION IS COPYRIGHT PROTECTED
Copyright © 2018 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 21 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

67 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

TECHNICAL SPECIFICATION

Ultrasonics – Measurements of electroacoustical parameters and acoustic output power of spherically curved transducers using the self-reciprocity method

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 17.140.50

ISBN 978-2-8322-5026-6

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	8
2 Normative references	8
3 Terms and definitions	8
4 Symbols	12
5 General	13
6 Requirements of the measurement system	14
6.1 Apparatus configuration	14
6.2 Measurement water tank.....	14
6.3 Fixturing, positioning and orientation systems.....	14
6.4 Reflector.....	14
6.5 Current monitor (probe)	14
6.6 Oscilloscope	15
6.7 Measurement hydrophone.....	15
7 Measurement of the effective half-aperture of the spherically curved transducer.....	15
7.1 Setup.....	15
7.2 Alignment and positioning of the hydrophone in the field.....	15
7.3 Measurements of the beamwidth and the effective half-aperture	15
7.4 Calculations of the focus half-angle and the effective area.....	16
8 Measurements of the electroacoustical parameters and the acoustic output power	16
8.1 Self-reciprocity method for transducer calibration.....	16
8.1.1 Experimental procedures	16
8.1.2 Criterion for checking the linearity of the focused field	16
8.1.3 Criterion for checking the reciprocity of the transducer	16
8.2 Calculations of the transmitting response to current (voltage) and voltage sensitivity.....	17
8.3 Calculations of the transmitting response at geometric focus to current (voltage).....	17
8.4 Calculation of the pulse-echo sensitivity level	18
8.5 Measurements of the radiation conductance and the mechanical quality factor Q_m	18
8.5.1 Calculations of the acoustic output power and the radiation conductance	18
8.5.2 Measurement of the frequency response of the radiation conductance	18
8.6 Measurement of the electroacoustic efficiency	18
8.6.1 Calculation of the electric input power	18
8.6.2 Calculation of the electroacoustic efficiency.....	18
8.7 Measurement of the electric impedance (admittance).....	19
9 Measurement uncertainty	19
Annex A (informative) Relation of the average amplitude reflection coefficient on a plane interface of water-stainless steel and the focus half-angle for a normally incident beam of a circular spherically curved transducer [1, 2]	20
Annex B (informative) Diffraction correction coefficient G_{sf} in the free-field self-reciprocity calibration method for circular spherically curved transducers in water neglecting attenuation [2, 3, 4].....	24

Annex C (informative) Calculation of the diffraction correction coefficient $G_{\text{sf}}(R/\lambda, \beta)$ in the free-field self-reciprocity calibration in a non-attenuating medium for a circular spherically curved transducer [2, 3, 4, 7]	26
Annex D (informative) Speed of sound and attenuation in water	28
D.1 General	28
D.2 Speed of sound for propagation in water	28
D.3 Acoustic attenuation coefficient for propagation in water	28
Annex E (informative) Principle of reciprocity calibration for spherically curved transducers [2, 3, 4]	29
E.1 Principle of reciprocity calibration for an ideal spherically focused field of a transducer	29
E.2 Principle of reciprocity calibration of a real spherically focused field of a transducer	30
E.3 Self-reciprocity calibration of a spherically curved transducer	30
Annex F (informative) Experimental arrangements	35
F.1 Experimental arrangement for determining the effective radius of a transducer [2, 3, 4, 13]	35
F.2 Experimental arrangement of the self-reciprocity calibration method for a spherically curved transducer [2, 3, 4, 13]	35
Annex G (informative) Relationships between the electroacoustical parameters used in this application [13]	37
G.1 Relations between the free-field transmitting response to voltage (current) and the voltage sensitivity with the radiation conductance	37
G.2 Relation of the radiation conductance and the electroacoustic efficiency	38
G.3 Relation of the transmitting response and voltage and acoustic output power	38
G.4 Relation of the pulse echo sensitivity and the radiation conductance	38
Annex H (informative) Evaluation and expression of uncertainty in the measurements of the radiation conductance	39
H.1 Executive standard	39
H.2 Evaluation of uncertainty in the measurement of the radiation conductance	39
H.2.1 Mathematical expression	39
H.2.2 Type A evaluation of standard uncertainty	40
H.2.3 Type B evaluation of standard uncertainty	40
H.2.4 Evaluation of the combined standard uncertainty for the radiation conductance	42
Bibliography	46
Figure A.1 – Relation curve of the amplitude reflection coefficient $r(\theta_i)$ on the interface of water-stainless steel for a plane wave with the incident angle θ_i	22
Figure A.2 – Average amplitude reflection coefficient $r_{\text{av}}(\beta)$ on the plane interface of water-stainless steel in the geometric focal plane of a spherically curved transducer vs. the focus half-angle β	23
Figure C.1 – Geometry of the concave radiating surface A of a spherically curved transducer and its virtual image surface A' for their symmetry of mirror-images about the geometric focal plane $(x, y, 0)$	26
Figure E.1 – Spherical coordinates	31
Figure E.2 Function $G_{\text{a}}(k \sin \theta)$, diffraction pattern $F_0(k \sin \theta)$ and $F_0^2(k \sin \theta)$ in the geometric focal plane [7]	32
Figure F.1 – Scheme of the measurement apparatus for determining the effective half-aperture of a transducer	35

Figure F.2 – Scheme of free-field self-reciprocity method applied to a spherically curved transducer 36

Table A.1 – Parameters used in calculation of the average amplitude reflection coefficient 21

Table A.2 – Amplitude reflection coefficient $r(\theta_i)$ on a plane interface of water-stainless steel for plane wave vs. the incident angle θ_i 21

Table A.3 – Average amplitude reflection coefficient $r_{av}(\beta)$ on plane interface of water-stainless steel in the geometric focal plane of a spherically curved transducer vs. the focus half-angle β 22

Table B.1 – Diffraction correction coefficients G_{sf} of a circular spherically curved transducer in the self-reciprocity calibration method [2, 3, 4] 24

Table D.1 – Dependence of speed of sound in water on temperature [5] 28


Table E.1 – G_a values dependent on $ka\sin\theta$ for $\beta \leq 45^\circ$ where $x = ka\sin\theta$ (according to O'Neil [7]) 32

Table E.2 – The $(R/\lambda)_{min}$ values dependent on β when $\theta_{max} \geq \theta_{Ga}$ and $\beta < 45^\circ$ for $G_a = 0,94; 0,95; 0,96; 0,97; 0,98; 0,99$ 33

Table H.1 – Type B evaluation of the standard uncertainties (SU) of input quantities in measurement 41

Table H.2 – Components of the standard uncertainty for the measurement of the radiation conductance using the self-reciprocity method 44

Table.H.3 – The measurement results and evaluated data of uncertainty for five transducers 45



INTERNATIONAL ELECTROTECHNICAL COMMISSION

**ULTRASONICS – MEASUREMENTS OF ELECTROACOUSTICAL
PARAMETERS AND ACOUSTIC OUTPUT POWER OF SPHERICALLY
CURVED TRANSDUCERS USING THE SELF-RECIPROCALITY METHOD**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. In exceptional circumstances, a technical committee may propose the publication of a technical specification when

- the required support cannot be obtained for the publication of an International Standard, despite repeated efforts, or
- the subject is still under technical development or where, for any other reason, there is the future but no immediate possibility of an agreement on an International Standard.

Technical specifications are subject to review within three years of publication to decide whether they can be transformed into International Standards.

IEC TS 62903, which is a Technical Specification, has been prepared by IEC technical committee 87: Ultrasonics.

The text of this technical specification is based on the following documents:

Enquiry draft	Report on voting
87/652/DTS	87/659/RVDTS

Full information on the voting for the approval of this technical specification can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- **terms defined in Clause 3: in bold type.**

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

Withdrawn

INTRODUCTION

An ultrasonic transducer is an important acoustic device that can act as a transmitter or a receiver in the applications of medical ultrasound, non-destructive testing, and ultrasonic materials processing. The performance of a transducer is a decisive factor that governs the device's range of applicability, efficiency and quality control in the manufacturing. The mechanisms, transmitting fields, performances, and measurement methods used for these transducers have been studied over the past few decades. However, the electroacoustical characterization and measurement methods applied for spherically curved transducers have not been defined in standard documents for either terms or protocols.

This document defines the relevant electroacoustical parameters for these devices and establishes the self-reciprocity measurement method for spherically curved concave focusing transducers.

Withdrawn