

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Radiation protection instrumentation – Measurement of personal dose equivalents for X, gamma, neutron and beta radiations – Active personal dosimeters

Instrumentation pour la radioprotection – Mesure des équivalents de dose individuels pour les rayonnements X, gamma, neutron et bêta – Dosimètres individuels actifs



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2024 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

IEC publications search - webstore.iec.ch/advsearchform

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

Electropedia - www.electropedia.org

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

Recherche de publications IEC -

webstore.iec.ch/advsearchform

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

IEC Just Published - webstore.iec.ch/justpublished

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

Service Clients - webstore.iec.ch/csc

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: sales@iec.ch.

IEC Products & Services Portal - products.iec.ch

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

Electropedia - www.electropedia.org

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.



IEC 61526

Edition 4.0 2024-03

INTERNATIONAL STANDARD

NORME INTERNATIONALE



Radiation protection instrumentation – Measurement of personal dose equivalents for X, gamma, neutron and beta radiations – Active personal dosimeters

Instrumentation pour la radioprotection – Mesure des équivalents de dose individuels pour les rayonnements X, gamma, neutron et bêta – Dosimètres individuels actifs

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

COMMISSION
ELECTROTECHNIQUE
INTERNATIONALE

ICS 13.280

ISBN 978-2-8322-8176-5

**Warning! Make sure that you obtained this publication from an authorized distributor.
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

CONTENTS

FOREWORD.....	7
INTRODUCTION.....	9
1 Scope.....	10
2 Normative references	11
3 Terms and definitions	13
4 Units and symbols	21
4.1 Units.....	21
4.2 Symbols.....	21
5 Mechanical characteristics.....	23
5.1 Case.....	23
5.2 Switches	23
6 General characteristics.....	24
6.1 Storage of dose information	24
6.2 Indication	24
6.3 Dosimeter markings	24
6.4 Retention of radioactive contamination.....	24
6.5 Ranges for dose and dose rate	24
6.6 Rated range of an influence quantity.....	24
6.7 Effect of radiation not intended to be measured and use of more than one dosimeter.....	25
6.8 Intrinsic and environmental background of the instrument	25
6.9 Dose or dose rate alarms	25
6.9.1 General	25
6.9.2 Dose alarms	25
6.9.3 Dose rate alarms	25
6.9.4 Alarm output.....	26
6.10 Indication of malfunction	26
6.11 Assignment of the dose value to the dosimeter	26
6.12 Algorithm to evaluate the indicated value	26
7 General test procedures	26
7.1 Nature of tests	26
7.2 Reference conditions and standard test conditions.....	27
7.3 Tests for influence quantities of type F.....	28
7.4 Tests for influence quantities of type S	28
7.5 Phantom for testing.....	28
7.6 Position of detector assembly for the purpose of testing.....	28
7.7 Position of dosimeter during use	28
7.8 Minimum rated range of influence quantity	29
7.9 Low dose rates	29
7.10 Statistical fluctuations	29
7.11 Production of reference radiation	29
8 Additivity of indicated value	29
8.1 Requirements	29
8.2 Method of test.....	30
8.2.1 General	30
8.2.2 Preparation of the test	30

8.2.3	Practical test	31
8.3	Interpretation of the results	31
9	Radiation performance requirements and tests	31
9.1	General.....	31
9.2	Consideration of the uncertainty of the conventional quantity value.....	32
9.3	Constancy of the dose response, dose rate dependence and statistical fluctuations	32
9.3.1	General	32
9.3.2	Requirements	32
9.3.3	Method of test using sources	37
9.3.4	Interpretation of the results of the test using sources	38
9.3.5	Method of test for photon dosimeters using natural radiation	39
9.3.6	Interpretation of the results of the test using natural radiation.....	39
9.4	Variation of the response due to photon radiation energy and angle of incidence	40
9.4.1	Measurement quantity $H_p(0,07)$ or $\dot{H}_p(0,07)$	40
9.4.2	Measurement quantity $H_p(3)$ or $\dot{H}_p(3)$	40
9.4.3	Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$	41
9.5	Variation of the response due to neutron radiation energy and angle of incidence	42
9.5.1	General	42
9.5.2	Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$	42
9.6	Variation of the response due to beta radiation energy and angle of incidence	43
9.6.1	General	43
9.6.2	Measurement quantity $H_p(0,07)$ or $\dot{H}_p(0,07)$	43
9.6.3	Measurement quantity $H_p(3)$ or $\dot{H}_p(3)$	44
9.6.4	Measurement quantity $H_p(10)$ or $\dot{H}_p(10)$	45
9.7	Over indication due to radiation incident from the side of an $H_p(10)$, $H_p(3)$ or $H_p(0,07)$ dosimeter	45
9.7.1	Requirements	45
9.7.2	Method of test.....	45
9.7.3	Interpretation of the results	46
9.8	Retention of dose reading	46
9.8.1	General	46
9.8.2	Requirements	46
9.8.3	Method of test and interpretation of the results	46
9.9	Overload characteristics	46
9.9.1	General	46
9.9.2	Requirements	47
9.9.3	Method of test and interpretation of the results	47
9.10	Alarm	48
9.10.1	General	48
9.10.2	Response time for dose rate indication and alarm.....	48
9.10.3	Accuracy of dose alarm	49
9.10.4	Accuracy of dose rate alarm	49
9.11	Model function	50
10	Electrical and environmental performance requirements and tests.....	50

10.1	General.....	50
10.2	Power supplies	50
10.2.1	General requirements	50
10.2.2	Specific primary batteries requirements	51
10.2.3	Specific secondary batteries requirements.....	51
10.2.4	Method of test and interpretation of the results (primary and secondary batteries)	51
10.3	Ambient temperature.....	52
10.3.1	Requirements	52
10.3.2	Method of test and interpretation of the results	53
10.4	Relative humidity	54
10.4.1	Requirements	54
10.4.2	Method of test and interpretation of the results	54
10.5	Atmospheric pressure	54
10.6	Sealing	54
10.7	Storage.....	54
10.8	Light exposure	54
10.8.1	General	54
10.8.2	Requirements	55
10.8.3	Interpretation of the results.....	55
10.9	Dose build-up, fading and self-irradiation (hybrid dosimeters).....	55
11	Electromagnetic performance requirements and tests.....	56
11.1	General.....	56
11.2	Electrostatic discharge.....	57
11.2.1	Requirements	57
11.2.2	Method of test and interpretation of the results	57
11.3	Radiated electromagnetic fields	58
11.3.1	Requirements	58
11.3.2	Method of test and interpretation of the results	58
11.4	Conducted disturbances induced by fast transients or bursts	58
11.4.1	Requirements	58
11.4.2	Method of test and interpretation of the results	58
11.5	Conducted disturbances induced by surges	59
11.5.1	Requirements	59
11.5.2	Method of test and interpretation of the results	59
11.6	Conducted disturbances induced by radio-frequencies.....	59
11.6.1	Requirements	59
11.6.2	Method of test and interpretation of the results	59
11.7	50 Hz/60 Hz magnetic field	59
11.7.1	Requirements	59
11.7.2	Method of test and interpretation of the results	59
11.8	Voltage dips and short interruptions.....	59
11.8.1	Requirements	59
11.8.2	Method of test and interpretation of the results	60
12	Mechanical performance, requirements and tests	60
12.1	General.....	60
12.2	Drop test.....	60
12.2.1	Requirements	60
12.2.2	Method of test and interpretation of the results	60

12.3	Vibration test	61
12.3.1	Requirements	61
12.3.2	Method of test and interpretation of the results	61
12.4	Microphonics test.....	61
12.4.1	Requirements	61
12.4.2	Method of test and interpretation of the results	61
13	Software.....	62
13.1	General.....	62
13.2	Design and structure of the software.....	63
13.2.1	Requirements	63
13.2.2	Method of test.....	63
13.3	Identification of the software	63
13.3.1	Requirements	63
13.3.2	Method of test.....	63
13.4	Authenticity of the software and the presentation of results.....	64
13.4.1	Requirements	64
13.4.2	Method of test.....	64
13.5	Alarm and stop of system operation under abnormal operating conditions.....	64
13.5.1	Requirements	64
13.5.2	Method of test.....	64
13.6	Control of instrument parameters	65
13.6.1	Requirements	65
13.6.2	Method of test.....	65
13.7	Storage of measurement data	65
13.7.1	Requirements	65
13.7.2	Method of test.....	65
13.8	Transmission of data.....	66
13.8.1	Requirements	66
13.8.2	Method of test.....	66
13.9	Hardware interfaces and software interfaces.....	66
13.9.1	Requirements	66
13.9.2	Method of test.....	67
13.10	Documentation for the software test.....	67
13.10.1	Requirements	67
13.10.2	Method of test.....	67
14	Uncertainty.....	68
15	Documentation	68
15.1	Type test report	68
15.2	Certificate	68
16	Operation and maintenance manual	69
Annex A (normative)	Statistical fluctuations	70
Annex B (normative)	Coverage intervals	72
B.1	General.....	72
B.2	Coverage interval for the mean, \bar{x}	73
B.3	Coverage interval for a combined quantity	73
Annex C (informative)	Causal connection between readout signals, indicated value and measured value	75

Annex D (informative) Procedure to determine the variation of the relative response due to radiation energy and angle of radiation incidence 76

Annex E (informative) Computational method of test for mixed irradiations 78

Bibliography..... 80

Figure B.1 – Test for coverage interval 72

Figure C.1 – Data evaluation in dosimetry systems 75

Figure E.1 – Flow chart of a computer program to perform tests according to 8.2 79

Table 1 – Symbols and abbreviated terms 21

Table 2 – Reference conditions and standard test conditions 27

Table 3 – Radiation characteristics of $H_p(0,07)$ dosimeters for X, gamma and beta radiation 33

Table 4 – Radiation characteristics of $H_p(3)$ dosimeters for X, gamma and beta radiation 34

Table 5 – Radiation characteristics of $H_p(10)$ dosimeters for X and gamma radiation 35

Table 6 – Radiation characteristics of $H_p(10)$ dosimeters for neutron radiation 36

Table 7 – Values of c_1 and c_2 for w different dose values and n indications for each dose value 39

Table 8 – Electrical and environmental characteristics of dosimeters 55

Table 9 – Electromagnetic disturbance characteristics of dosimeters 57

Table 10 – Mechanical disturbances characteristics of dosimeters 60

Table A.1 – Number of instrument readings required to detect true differences (95 % level of confidence) between two sets of instrument readings on the same instrument 71

Table B.1 – Student’s t -value for a double sided 95 % coverage interval..... 73

Table E.1 – Example of dosimeter response table and range limits..... 78

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**RADIATION PROTECTION INSTRUMENTATION –
MEASUREMENT OF PERSONAL DOSE EQUIVALENTS FOR X,
GAMMA, NEUTRON AND BETA RADIATIONS –
ACTIVE PERSONAL DOSEMETERS**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61526 has been prepared by subcommittee 45B: Radiation protection instrumentation, of IEC technical committee 45: Nuclear instrumentation. It is an International Standard.

This fourth edition cancels and replaces the third edition published in 2010. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) Modification of the title;
- b) Inclusion of the measurement quantity for the dose in the lens of the eye, $H_p(3)$;
- c) Inclusion of measurement quantity for dose in the skin and extremities, $H_p(0,07)$;

- d) Inclusion of dosimeters between active and passive: "hybrid dosimeters";
- e) Inclusion of software requirements;
- f) Harmonization of requirements for linearity to IEC 62387;
- g) Revised neutron energy response requirements.

The text of this International Standard is based on the following documents:

Draft	Report on voting
45B/1047/FDIS	45B/1049/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at www.iec.ch/members_experts/refdocs. The main document types developed by IEC are described in greater detail at www.iec.ch/publications.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under webstore.iec.ch in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This document applies to active, (direct reading or hybrid) personal dosimeters and monitors used for measuring personal dose equivalents $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, for X, gamma, neutron, and beta radiations.

For personal dose equivalent $H_p(10)$ and for X and gamma radiation, two minimum rated ranges for the photon energy are given. The first from 20 keV to 150 keV is for workplaces where low energy X-rays are used, e.g., in diagnostic medicine, the second from 80 keV to 1,25 MeV is for workplaces where high energy X-rays and/or gamma sources are used, e.g., in industry. For neutron radiation the minimum rated range of neutron energy is from 0,025 eV (thermal neutrons) to 10 MeV. The rated ranges can be extended to all energies covered by the respective standards for reference radiation fields.

For personal dose equivalent $H_p(3)$ and for X and gamma radiation, a minimum rated range for photon energy from 30 keV to 250 keV is given. For personal dose equivalent $H_p(0,07)$ a range of 30 keV to 1250 keV or, for workplaces where low energy X-rays are used, 20 keV to 150 keV, is given. For beta radiation for both quantities, the minimal rated range is from 0,24 MeV to 0,8 MeV (mean beta particle energy). The rated ranges can be extended to all energies covered by the respective standards for reference radiation fields.

In some applications, for example, at a nuclear reactor installation where 6 MeV photon radiation is present, measurement of personal dose equivalent (rate) $H_p(10)$ for photon energies up to 10 MeV should be required. In some other applications, measurement of $H_p(10)$ down to 10 keV should be required.

For personal dosimeters, requirements for measuring the dose quantities $H_p(10)$, $H_p(3)$, and $H_p(0,07)$, and for monitoring of the respective dose rate quantities are given. The measurement of these dose rate quantities is an option for personal dosimeters.

Establishments in some countries may be permitted to use this type of personal dosimeter as the dosimeter to provide the dose of record by an approved dosimetry service.